

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872
Instrument ID: BNA_M Calibration Date(s): 04/28/2025 04/28/2025
Calibration Time(s): 12:30 17:04

LAB FILE ID:		RRF2.5 = BM050024.D		RRF005 = BM050025.D		RRF010 = BM050026.D			
		RRF020 = BM050027.D		RRF040 = BM050028.D		RRF050 = BM050029.D			
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD	
2-Fluorophenol		1.074	1.079	1.107	1.208	1.211	1.142	5.0	
Phenol-d6		1.290	1.318	1.398	1.543	1.537	1.436	7.1	
Nitrobenzene-d5		0.386	0.374	0.397	0.434	0.433	0.406	5.7	
2-Fluorobiphenyl		1.588	1.555	1.632	1.807	1.822	1.691	6.2	
2,4,6-Tribromophenol		0.253	0.249	0.281	0.318	0.329	0.296	11.6	
Atrazine		0.196	0.198	0.214	0.236	0.242	0.222	8.5	
Terphenyl-d14		1.176	1.189	1.332	1.524	1.541	1.352	11.6	