

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872
Instrument ID: BNA_P Calibration Date(s): 04/14/2025 04/14/2025
Calibration Time(s): 11:06 17:13

LAB FILE ID:		RRF2.5 = BP024275.D		RRF005 = BP024276.D		RRF010 = BP024277.D		
		RRF020 = BP024278.D		RRF040 = BP024279.D		RRF050 = BP024280.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.210	4.9
Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.656	6.7
Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.351	4.8
2-Fluorobiphenyl		1.332	1.334	1.359	1.366	1.263	1.306	4.1
2,4,6-Tribromophenol		0.235	0.256	0.278	0.289	0.289	0.277	8.3
Atrazine		0.176	0.167	0.134	0.121	0.162	0.152	15.6
Terphenyl-d14		0.967	0.998	1.055	1.032	1.005	0.992	4.4