

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872

Instrument ID: BNA_P Calibration Date/Time: 05/08/2025 14:45

Lab File ID: BP024568.D Init. Calib. Date(s): 04/14/2025 04/14/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:06 17:13

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.210	1.238		2.3	
Phenol-d6	1.656	1.642		-0.8	
Nitrobenzene-d5	0.351	0.357		1.7	
2-Fluorobiphenyl	1.306	1.381		5.7	
2,4,6-Tribromophenol	0.277	0.353		27.4	
Atrazine	0.152	0.143		-5.9	
Terphenyl-d14	0.992	1.078		8.7	

All other compounds must meet a minimum RRF of 0.010.