

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872

Instrument ID: BNA_P Calibration Date/Time: 05/14/2025 11:35

Lab File ID: BP024625.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:41 15:26

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.169	1.128		-3.5	
Phenol-d6	1.492	1.457		-2.3	
Nitrobenzene-d5	0.396	0.391		-1.3	
2-Fluorobiphenyl	1.527	1.435		-6.0	
2,4,6-Tribromophenol	0.339	0.349		3.0	
Atrazine	0.225	0.219		-2.7	
Terphenyl-d14	1.166	1.115		-4.4	

All other compounds must meet a minimum RRF of 0.010.