

Surrogate Summary

SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167904BL	PB167904BL	Nitrobenzene-d5	100	88.3	88		18	107
		2-Fluorobiphenyl	100	95.4	95		20	109
		Terphenyl-d14	100	95.5	96		10	105
PB167904BS	PB167904BS	Nitrobenzene-d5	100	88.0	88		18	107
		2-Fluorobiphenyl	100	91.1	91		20	109
		Terphenyl-d14	100	102	102		10	105
Q1872-12	HW0425-PT-TRIAZINE-SOIL	Nitrobenzene-d5	100	71.2	71		18	107
		2-Fluorobiphenyl	100	73.9	74		20	109
		Terphenyl-d14	100	92.8	93		10	105
Q1964-01MS	MH-LLMS	Nitrobenzene-d5	100	62.7	63		18	107
		2-Fluorobiphenyl	100	55.9	56		20	109
		Terphenyl-d14	100	51.3	51		10	105
Q1964-01MSD	MH-LLMSD	Nitrobenzene-d5	100	60.5	61		18	107
		2-Fluorobiphenyl	100	53.4	53		20	109
		Terphenyl-d14	100	47.9	48		10	105