

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/13/2025

Lab File ID: BN036992.D Time Analyzed: 10:39

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1673	7.618	4334	10.40	2424	14.27
UPPER LIMIT	3346	8.118	8668	10.904	4848	14.766
LOWER LIMIT	836.5	7.118	2167	9.904	1212	13.766
EPA SAMPLE NO.						
01 HW0425-PT-TRIAZINE-SOIL	2166	7.62	6095	10.40	3427	14.27

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.