

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM

SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BN036922.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:56

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	67.4
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	58.8
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	54.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	8.4
442	Greater than 50% of mass 198	48.1
443	15.0 - 24.0% of mass 442	9.3 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036923.D	04/28/2025	11:35
SSTDICC0.2	SSTDICC0.2	BN036924.D	04/28/2025	12:11
SSTDICCC0.4	SSTDICCC0.4	BN036925.D	04/28/2025	12:47
SSTDICC0.8	SSTDICC0.8	BN036926.D	04/28/2025	13:24
SSTDICC1.6	SSTDICC1.6	BN036927.D	04/28/2025	14:00
SSTDICC3.2	SSTDICC3.2	BN036928.D	04/28/2025	14:36
SSTDICC5.0	SSTDICC5.0	BN036929.D	04/28/2025	15:12