

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022046.D
 Acq On : 29 Apr 2025 09:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.426	0.371	12.9	97	0.00
3 P	Chloromethane	0.607	0.532	12.4	91	0.00
4 C	Vinyl Chloride	0.745	0.666	10.6#	94	0.00
5 T	Bromomethane	0.642	0.514	19.9	89	0.00
6 T	Chloroethane	0.508	0.429	15.6	89	0.00
7 T	Trichlorofluoromethane	0.983	0.856	12.9	92	0.00
8 T	Diethyl Ether	0.256	0.216	15.6	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.466	12.6	98	0.00
10 T	Methyl Iodide	0.617	0.574	7.0	94	0.00
11 T	Tert butyl alcohol	0.025	0.023	8.0	89	0.00
12 CM	1,1-Dichloroethene	0.497	0.442	11.1#	96	0.00
13 T	Acrolein	0.046	0.043	6.5	103	0.00
14 T	Allyl chloride	0.579	0.501	13.5	93	0.00
15 T	Acrylonitrile	0.094	0.084	10.6	92	0.00
16 T	Acetone	0.070	0.065	7.1	106	0.00
17 T	Carbon Disulfide	1.585	1.366	13.8	94	0.00
18 T	Methyl Acetate	0.216	0.209	3.2	96	0.00
19 T	Methyl tert-butyl Ether	1.113	1.008	9.4	91	0.00
20 T	Methylene Chloride	0.584	0.468	19.9	90	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.489	12.2	94	0.01
22 T	Diisopropyl ether	1.288	1.156	10.2	92	0.00
23 T	Vinyl Acetate	0.694	0.630	9.2	90	0.00
24 P	1,1-Dichloroethane	0.893	0.776	13.1	93	0.00
25 T	2-Butanone	0.105	0.098	6.7	95	0.00
26 T	2,2-Dichloropropane	0.782	0.709	9.3	98	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.558	10.3	94	0.00
28 T	Bromochloromethane	0.347	0.322	7.2	96	0.00
29 T	Tetrahydrofuran	0.068	0.063	7.4	89	0.00
30 C	Chloroform	0.990	0.877	11.4#	94	0.00
31 T	Cyclohexane	0.763	0.661	13.4	96	0.00
32 T	1,1,1-Trichloroethane	0.896	0.789	11.9	95	0.00
33 S	1,2-Dichloroethane-d4	0.462	0.402	13.0	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.330	0.314	4.8	92	0.00
36 T	1,1-Dichloropropene	0.444	0.412	7.2	94	0.00
37 T	Ethyl Acetate	0.159	0.152	4.4	91	0.00
38 T	Carbon Tetrachloride	0.530	0.509	4.0	97	0.00
39 T	Methylcyclohexane	0.558	0.529	5.2	93	0.00
40 TM	Benzene	1.387	1.301	6.2	95	0.00
41 T	Methacrylonitrile	0.087	0.072	17.2	76	0.00
42 TM	1,2-Dichloroethane	0.343	0.319	7.0	92	0.00
43 T	Isopropyl Acetate	0.306	0.284	7.2	88	0.00
44 TM	Trichloroethene	0.384	0.360	6.3	95	0.00
45 C	1,2-Dichloropropane	0.307	0.287	6.5#	93	0.00
46 T	Dibromomethane	0.192	0.181	5.7	94	0.00
47 T	Bromodichloromethane	0.480	0.457	4.8	94	0.00
48 T	Methyl methacrylate	0.141	0.139	1.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022046.D
 Acq On : 29 Apr 2025 09:28
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 01:32:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	98	0.00
50 S	Toluene-d8	1.245	1.202	3.5	92	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.161	-0.6	91	0.00
52 CM	Toluene	0.898	0.870	3.1#	95	0.00
53 T	t-1,3-Dichloropropene	0.411	0.393	4.4	92	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.462	5.5	93	0.00
55 T	1,1,2-Trichloroethane	0.256	0.247	3.5	94	0.00
56 T	Ethyl methacrylate	0.293	0.293	0.0	90	0.00
57 T	1,3-Dichloropropane	0.411	0.395	3.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.152	-6.3	92	0.00
59 T	2-Hexanone	0.106	0.109	-2.8	92	0.00
60 T	Dibromochloromethane	0.355	0.349	1.7	96	0.00
61 T	1,2-Dibromoethane	0.242	0.236	2.5	95	0.00
62 S	4-Bromofluorobenzene	0.419	0.404	3.6	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.472	0.449	4.9	98	0.00
65 PM	Chlorobenzene	1.118	1.034	7.5	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.367	6.9	94	0.00
67 C	Ethyl Benzene	1.829	1.744	4.6#	94	0.00
68 T	m/p-Xylenes	0.728	0.712	2.2	96	0.00
69 T	o-Xylene	0.671	0.649	3.3	93	0.00
70 T	Styrene	1.126	1.110	1.4	94	0.00
71 P	Bromoform	0.229	0.224	2.2	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.341	3.152	5.7	95	0.00
74 T	N-amyl acetate	0.574	0.524	8.7	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.563	0.517	8.2	94	0.00
76 T	1,2,3-Trichloropropane	0.428	0.403	5.8	100	0.00
77 T	Bromobenzene	0.844	0.794	5.9	96	0.00
78 T	n-propylbenzene	3.978	3.763	5.4	95	0.00
79 T	2-Chlorotoluene	2.314	2.139	7.6	94	0.00
80 T	1,3,5-Trimethylbenzene	2.752	2.633	4.3	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.178	0.163	8.4	92	0.00
82 T	4-Chlorotoluene	2.405	2.260	6.0	95	0.00
83 T	tert-Butylbenzene	2.468	2.334	5.4	94	0.00
84 T	1,2,4-Trimethylbenzene	2.753	2.637	4.2	95	0.00
85 T	sec-Butylbenzene	3.612	3.431	5.0	95	0.00
86 T	p-Isopropyltoluene	3.067	2.964	3.4	96	0.00
87 T	1,3-Dichlorobenzene	1.714	1.588	7.4	96	0.00
88 T	1,4-Dichlorobenzene	1.698	1.573	7.4	97	0.00
89 T	n-Butylbenzene	2.742	2.563	6.5	93	0.00
90 T	Hexachloroethane	0.674	0.613	9.1	96	0.00
91 T	1,2-Dichlorobenzene	1.495	1.376	8.0	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.080	11.1	95	0.00
93 T	1,2,4-Trichlorobenzene	0.847	0.803	5.2	98	0.00
94 T	Hexachlorobutadiene	0.521	0.484	7.1	101	0.00
95 T	Naphthalene	1.408	1.372	2.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
Data File : VY022046.D
Acq On : 29 Apr 2025 09:28
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 30 01:32:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.731	0.700	4.2	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6