

Surrogate Summary

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1872-23	HW0425-PT-VOA-SOIL	1,2-Dichloroethane-d4	50	48.1	96	63	155
		Dibromofluoromethane	50	48.5	97	70	134
		Toluene-d8	50	50.8	102	74	123
		4-Bromofluorobenzene	50	47.4	95	38	136
Q1872-23DL	HW0425-PT-VOA-SOILDL	1,2-Dichloroethane-d4	50	50.8	102	63	155
		Dibromofluoromethane	50	49.7	99	70	134
		Toluene-d8	50	51.3	103	74	123
		4-Bromofluorobenzene	50	47.2	94	38	136
VY0428SBL01	VY0428SBL01	1,2-Dichloroethane-d4	50	51.3	103	63	155
		Dibromofluoromethane	50	50.8	102	70	134
		Toluene-d8	50	47.1	94	74	123
		4-Bromofluorobenzene	50	57.4	115	38	136
VY0428SBS01	VY0428SBS01	1,2-Dichloroethane-d4	50	49.5	99	63	155
		Dibromofluoromethane	50	50.8	101	70	134
		Toluene-d8	50	51.9	104	74	123
		4-Bromofluorobenzene	50	50.6	101	38	136
VY0428SBSD01	VY0428SBSD01	1,2-Dichloroethane-d4	50	48.0	96	63	155
		Dibromofluoromethane	50	50.0	100	70	134
		Toluene-d8	50	50.3	101	74	123
		4-Bromofluorobenzene	50	48.8	98	38	136
VY0429SBL01	VY0429SBL01	1,2-Dichloroethane-d4	50	50.3	101	63	155
		Dibromofluoromethane	50	49.9	100	70	134
		Toluene-d8	50	47.6	95	74	123
		4-Bromofluorobenzene	50	56.5	113	38	136
VY0429SBS01	VY0429SBS01	1,2-Dichloroethane-d4	50	48.4	97	63	155
		Dibromofluoromethane	50	50.2	100	70	134
		Toluene-d8	50	50.9	102	74	123
		4-Bromofluorobenzene	50	49.6	99	38	136