

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ALLI03
 Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872
 Lab File ID: VY022021.D Date Analyzed: 04/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	238558	13.346				
UPPER LIMIT	477116	13.846				
LOWER LIMIT	119279	12.846				
EPA SAMPLE NO.						
HW0425-PT-VOA-SOIL	221687	13.35				
VY0428SBL01	183874	13.35				
VY0428SBS01	223411	13.35				
VY0428SBSD01	228742	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.