

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872
Lab File ID: VY022020.D BFB Injection Date: 04/28/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:29
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.1
75	30.0 - 60.0% of mass 95	47.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1.6 (1.8) 1
174	50.0 - 100.0% of mass 95	88.6
175	5.0 - 9.0% of mass 174	6.7 (7.6) 1
176	95.0 - 101.0% of mass 174	85.9 (96.9) 1
177	5.0 - 9.0% of mass 176	5.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022021.D	04/28/2025	09:59
VY0428SBL01	VY0428SBL01	VY022022.D	04/28/2025	11:08
VY0428SBS01	VY0428SBS01	VY022023.D	04/28/2025	11:40
VY0428SBSD01	VY0428SBSD01	VY022024.D	04/28/2025	12:02
HW0425-PT-VOA-SOIL	Q1872-23	VY022025.D	04/28/2025	12:41