

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1875 SAS No.: Q1875 SDG NO.: Q1875

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/28/2025

Lab File ID: BP024435.D Time Analyzed: 13:34

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161036	7.716	671932	10.49	431586	14.34
UPPER LIMIT	322072	8.216	1343860	10.987	863172	14.839
LOWER LIMIT	80518	7.216	335966	9.987	215793	13.839
EPA SAMPLE NO.						
01 PB167739BL	161610	7.72	627258	10.49	404300	14.35
02 PB167739BS	174066	7.72	694511	10.49	444446	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.