

Surrogate Summary

SDG No.: Q1875

Client: PSEG

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1875-02	AUD-25-0053	1,2-Dichloroethane-d4	50	54.8	110	74	125
		Dibromofluoromethane	50	50.9	102	75	124
		Toluene-d8	50	52.1	104	86	113
		4-Bromofluorobenzene	50	53.3	107	77	121

Surrogate Summary

SDG No.: Q1875

Client: PSEG

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0428WBL01	VX0428WBL01	1,2-Dichloroethane-d4	50	54.6	109	74	125
		Dibromofluoromethane	50	51.5	103	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VX0428WBS01	VX0428WBS01	1,2-Dichloroethane-d4	50	54.4	109	74	125
		Dibromofluoromethane	50	55.5	111	75	124
		Toluene-d8	50	52.5	105	86	113
		4-Bromofluorobenzene	50	55.1	110	77	121
VX0428WBSD0	VX0428WBSD01	1,2-Dichloroethane-d4	50	52.1	104	74	125
		Dibromofluoromethane	50	53.5	107	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	54.1	108	77	121

