

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1875

Client: PSEG

Analytical Method: SW8260-Low

Datafile : VX045947.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0428WBS01	Vinyl chloride	20	17.7	ug/L	89			65	117	
	1,1-Dichloroethene	20	19.4	ug/L	97			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	21.9	ug/L	110			77	113	
	Chloroform	20	21.4	ug/L	107			79	113	
	Benzene	20	20.2	ug/L	101			82	109	
	1,2-Dichloroethane	20	21.9	ug/L	110			80	115	
	Trichloroethene	20	19.6	ug/L	98			77	113	
	Tetrachloroethene	20	20.3	ug/L	102			67	123	
	Chlorobenzene	20	21.2	ug/L	106			82	109	