

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1875

Client: PSEG

Analytical Method: SW8260-Low

Datafile : VX045948.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0428WBSD01	Vinyl chloride	20	16.5	ug/L	83	7		65	117	20
	1,1-Dichloroethene	20	17.8	ug/L	89	9		74	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	20.7	ug/L	104	6		77	113	20
	Chloroform	20	20.5	ug/L	103	4		79	113	20
	Benzene	20	19.3	ug/L	97	4		82	109	20
	1,2-Dichloroethane	20	21.6	ug/L	108	2		80	115	20
	Trichloroethene	20	19.2	ug/L	96	2		77	113	20
	Tetrachloroethene	20	19.8	ug/L	99	3		67	123	20
	Chlorobenzene	20	20.5	ug/L	103	3		82	109	20