

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135555

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	04/25/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	04/25/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	04/25/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135559

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.94	1	94	90-110	04/25/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	04/25/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.99	1	99	90-110	04/25/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.93	1	93	90-110	04/25/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	04/25/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135562

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TKN		mg/L	4.7	5	94	90-110	04/25/2025
Sample ID:	CCV1						
TKN		mg/L	4.8	5	96	90-110	04/25/2025
Sample ID:	CCV2						
TKN		mg/L	5.1	5	102	90-110	04/25/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135563

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Phenolics	mg/L	1.1	1	110	90-110	04/25/2025
Sample ID: CCV1 Phenolics	mg/L	1	1	100	90-110	04/25/2025
Sample ID: CCV2 Phenolics	mg/L	1	1	100	90-110	04/25/2025