

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135559

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	04/25/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	04/25/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	04/25/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	04/25/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	04/25/2025

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135562

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TKN	mg/L	0.13	0.2500	J	0.11	0.5	04/25/2025
Sample ID: CCB1 TKN	mg/L	0.13	0.2500	J	0.11	0.5	04/25/2025
Sample ID: CCB2 TKN	mg/L	0.15	0.2500	J	0.11	0.5	04/25/2025

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

RunNo.: LB135563

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	04/25/2025
Sample ID: CCB1 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	04/25/2025
Sample ID: CCB2 Phenolics	mg/L	< 0.0250	0.0250	U	0.015	0.05	04/25/2025