

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

Run No.: LB135557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135557BS							
Acidity	mg/L	2500	2230		89	1	80-120	04/25/2025

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

Run No.: LB135561

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135561BS							
Alkalinity	mg/L	50	46.1		92	1	80-120	04/25/2025

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

Run No.: LB135559

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167733BS							
Ammonia as N	mg/L	1	0.93		93	1	90-110	04/25/2025

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

Run No.: LB135562

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167734BS							
TKN	mg/L	5	4.80		96	1	90-110	04/25/2025

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q1879

Project: Tank Farm - Acid Analysis

Run No.: LB135563

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167735BS							
Phenolics	mg/L	1	1.10		110	1	80-120	04/25/2025