

LB135560


7470A

INSTRUMENT ID: CV1

Linear

$\mu$  Abs.:

48619



A= 0.0000e+000

B= 2.0119e-004

C= -7.9072e-002

Rho= 0.9983968

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/A
0.0	0.000	-0.057	-0.057	109	0.000	109	0				-
0.2	0.200	0.191	-0.009	1341	0.0 %	1341					-5
2.5	2.500	2.341	-0.159	12027	0.0 %	12027					-6
5.0	5.000	5.325	0.325	26858	0.0 %	26858					-7
7.5	7.500	7.698	0.198	38656	0.0 %	38656					-8
10.0	10.000	9.703	-0.297	48619	0.0 %	48619					-3

# LB135560 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Apr 2025 12:44:40

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	109	-	0.0000	7470A	PPB	25 Apr 2025 13:01:33	S	Std
0.2 - 1	502	1341	-	0.2000	7470A	PPB	25 Apr 2025 13:03:50	S	Std
2.5 - 1		12027	-	2.5000	7470A	PPB	25 Apr 2025 13:06:07	S	Std
5.0 - 1	2.5	26858	-	5.0000	7470A	PPB	25 Apr 2025 13:08:24	S	Std
7.5 - 1	5.5	38656	-	7.5000	7470A	PPB	25 Apr 2025 13:10:42	S	Std
10.0 - 1	32.5	48619	-	10.0000	7470A	PPB	25 Apr 2025 13:13:00	S	Std
ICV102 - 1	ICV102	19618	3.8680	-	7470A	PPB	25 Apr 2025 13:19:06	U	SMPL
ICB102 - 1	ICB102	246	-0.0296	-	7470A	PPB	25 Apr 2025 13:21:21	U	SMPL
CCV29 - 1	CCV29	26783	5.3096	-	7470A	PPB	25 Apr 2025 13:23:36	U	SMPL
CCB29 - 1	CCB29	168	-0.0453	-	7470A	PPB	25 Apr 2025 13:25:53	U	SMPL
CRA - 1	CRA	1308	0.1841	-	7470A	PPB	25 Apr 2025 13:28:10	U	SMPL
HighStd - 1	HighStd	49023	9.7841	-	7470A	PPB	25 Apr 2025 13:30:25	U	SMPL
ChkStd - 1	ChkStd	35866	7.1370	-	7470A	PPB	25 Apr 2025 13:32:40	U	SMPL
PB167749BL - 1	PBW	165	-0.0459	-	7470A	PPB	25 Apr 2025 13:34:59	U	SMPL
PB167749BS - 1	LCSW	18566	3.6563	-	7470A	PPB	25 Apr 2025 13:37:17	U	SMPL
Q1879-01 - 1	STEM-31-SULFURIC-ACID	13312	2.5992	-	7470A	PPB	25 Apr 2025 13:39:34	U	SMPL
Q1879-02 - 1	STEM-32-SULFURIC-ACID	382	-0.0022	-	7470A	PPB	25 Apr 2025 13:41:52	U	SMPL
Q1879-03 - 1	STEM-46-SULFURIC-ACID	628	0.0473	-	7470A	PPB	25 Apr 2025 13:44:11	U	SMPL
Q1879-04 - 1	ACID-MIXTURE-TF-CONTAINMENT	7322	1.3941	-	7470A	PPB	25 Apr 2025 13:46:29	U	SMPL
CCV30 - 1	CCV30	27287	5.4109	-	7470A	PPB	25 Apr 2025 13:58:58	U	SMPL
CCB30 - 1	CCB30	37	-0.0716	-	7470A	PPB	25 Apr 2025 14:01:15	U	SMPL