



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: METEM A GE POWER Business **SDG No.:** Q1879
Contract: METE01 **Lab Code:** CHEM **Case No.:** Q1879 **SAS No.:** Q1879
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	92	40 - 160	CV	04/25/2025	13:28	LB135560
CRI01	Aluminum	94.9	100	95	40 - 160	P	04/25/2025	14:01	LB135566
	Antimony	53.0	50.0	106	40 - 160	P	04/25/2025	14:01	LB135566
	Arsenic	21.0	20.0	105	40 - 160	P	04/25/2025	14:01	LB135566
	Barium	88.5	100	88	40 - 160	P	04/25/2025	14:01	LB135566
	Beryllium	6.00	6.0	100	40 - 160	P	04/25/2025	14:01	LB135566
	Cadmium	6.11	6.0	102	40 - 160	P	04/25/2025	14:01	LB135566
	Calcium	1910	2000	95	40 - 160	P	04/25/2025	14:01	LB135566
	Chromium	9.52	10.0	95	40 - 160	P	04/25/2025	14:01	LB135566
	Cobalt	29.2	30.0	97	40 - 160	P	04/25/2025	14:01	LB135566
	Copper	22.4	20.0	112	40 - 160	P	04/25/2025	14:01	LB135566
	Iron	87.9	100	88	40 - 160	P	04/25/2025	14:01	LB135566
	Lead	10.3	12.0	86	40 - 160	P	04/25/2025	14:01	LB135566
	Magnesium	2050	2000	102	40 - 160	P	04/25/2025	14:01	LB135566
	Manganese	19.9	20.0	100	40 - 160	P	04/25/2025	14:01	LB135566
	Nickel	39.0	40.0	97	40 - 160	P	04/25/2025	14:01	LB135566
	Potassium	1860	2000	93	40 - 160	P	04/25/2025	14:01	LB135566
	Selenium	15.5	20.0	77	40 - 160	P	04/25/2025	14:01	LB135566
	Silver	9.96	10.0	100	40 - 160	P	04/25/2025	14:01	LB135566
	Sodium	1850	2000	93	40 - 160	P	04/25/2025	14:01	LB135566
	Thallium	42.8	40.0	107	40 - 160	P	04/25/2025	14:01	LB135566
	Vanadium	38.6	40.0	97	40 - 160	P	04/25/2025	14:01	LB135566
	Zinc	41.0	40.0	102	40 - 160	P	04/25/2025	14:01	LB135566