

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1879

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167749BL		WATER		Batch Number:	PB167749		Prep Date:	04/25/2025	
	Mercury	0.20	<0.20	U	0.20	CV	04/25/2025	13:34	LB135560

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1879

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167747BL	WATER			Batch Number:	PB167747		Prep Date:	04/25/2025	
	Aluminum	50.0	<50.0	U	50.0	P	04/25/2025	15:55	LB135566
	Antimony	25.0	<25.0	U	25.0	P	04/25/2025	15:55	LB135566
	Arsenic	10.0	<10.0	U	10.0	P	04/25/2025	15:55	LB135566
	Barium	50.0	<50.0	U	50.0	P	04/25/2025	15:55	LB135566
	Beryllium	3.00	<3.00	U	3.00	P	04/25/2025	15:55	LB135566
	Cadmium	3.00	<3.00	U	3.00	P	04/25/2025	15:55	LB135566
	Calcium	1000	<1000	U	1000	P	04/25/2025	15:55	LB135566
	Chromium	5.00	<5.00	U	5.00	P	04/25/2025	15:55	LB135566
	Cobalt	15.0	<15.0	U	15.0	P	04/25/2025	15:55	LB135566
	Copper	10.0	<10.0	U	10.0	P	04/25/2025	15:55	LB135566
	Iron	50.0	<50.0	U	50.0	P	04/25/2025	15:55	LB135566
	Lead	6.00	<6.00	U	6.00	P	04/25/2025	15:55	LB135566
	Magnesium	1000	<1000	U	1000	P	04/25/2025	15:55	LB135566
	Manganese	10.0	<10.0	U	10.0	P	04/25/2025	15:55	LB135566
	Nickel	2.05	<20.0	J	20.0	P	04/25/2025	15:55	LB135566
	Potassium	1000	<1000	U	1000	P	04/25/2025	15:55	LB135566
	Selenium	10.0	<10.0	U	10.0	P	04/25/2025	15:55	LB135566
	Silver	5.00	<5.00	U	5.00	P	04/25/2025	15:55	LB135566
	Sodium	1000	<1000	U	1000	P	04/25/2025	15:55	LB135566
	Thallium	20.0	<20.0	U	20.0	P	04/25/2025	15:55	LB135566
	Vanadium	20.0	<20.0	U	20.0	P	04/25/2025	15:55	LB135566
	Zinc	20.0	<20.0	U	20.0	P	04/25/2025	15:55	LB135566