

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1879

Contract: METE01

Lab Code: CHEM

Case No.: Q1879

SAS No.: Q1879

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167747BS							
Aluminum	ug/L	1000	937		94	80 - 120	P
Antimony	ug/L	400	412		103	80 - 120	P
Arsenic	ug/L	400	399		100	80 - 120	P
Barium	ug/L	100	89.4		89	80 - 120	P
Beryllium	ug/L	100	94.8		95	80 - 120	P
Cadmium	ug/L	100	97.7		98	80 - 120	P
Calcium	ug/L	500	484	J	97	80 - 120	P
Chromium	ug/L	200	192		96	80 - 120	P
Cobalt	ug/L	100	95.2		95	80 - 120	P
Copper	ug/L	150	150		100	80 - 120	P
Iron	ug/L	1500	1400		93	80 - 120	P
Lead	ug/L	500	484		97	80 - 120	P
Magnesium	ug/L	1000	925	J	92	80 - 120	P
Manganese	ug/L	100	96.2		96	80 - 120	P
Nickel	ug/L	250	240		96	80 - 120	P
Potassium	ug/L	5000	4490		90	80 - 120	P
Selenium	ug/L	1000	1010		101	80 - 120	P
Silver	ug/L	37.5	36.1		96	80 - 120	P
Sodium	ug/L	1500	1380		92	80 - 120	P
Thallium	ug/L	1000	995		100	80 - 120	P
Vanadium	ug/L	150	140		93	80 - 120	P
Zinc	ug/L	100	97.7		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1879

Contract: METE01

Lab Code: CHEM

Case No.: Q1879

SAS No.: Q1879

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167749BS Mercury	ug/L	4.0	3.66		92	80 - 120	CV