



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech, EMI

Project Name: R36900

Project # N/A

Chemtech Project # Q1880

Test Name: TO-15

A. Number of Samples and Date of Receipt:

11 Air samples were received on 04/24/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:
TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for {Q1880-10DUP} with File ID: VL042430.D met criteria except for m/p-Xylene[200%], due to difference in results of original and DUP.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

Samples EBR-P11-SC-SS01-20250421, EBR-P16-SC-SS01-20250422 were diluted due to high concentrations.

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The Sample # EBR-P02-SL-IA02-20250423, EBR-P08-SC-IA01-20250421, EBR-P08-SC-SS01-20250421, EBR-P02-SC-DUP01-20250423, EBR-P02-SC-IA01-20250423 and EBR-P16-SC-AA01-20250422 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The Manual Integrations are performed for the followings.

Manual Integration Report			
Sequence	VL041725	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICCC010	VL042349.D	m/p-Xylene	SAM	4/18/2025 4:19:07 PM	MMDado da	4/18/2025 10:45:20 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042350.D	1,4-Dioxane	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042350.D	cis-1,3-Dichloropropene	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042350.D	Ethanol	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly
VSTDICCC002	VL042350.D	Heptane	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly

							ly
VSTDIC002	VL042350.D	m/p-Xylene	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly
VSTDIC002	VL042350.D	Trichloroethene	SAM	4/18/2025 4:19:12 PM	MMDado da	4/18/2025 10:45:17 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	1,1,2-Trichloroethane	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	1,3-Butadiene	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	1,4-Dioxane	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	2,2,4-Trimethylpentane	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	cis-1,3-Dichloropropene	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software

							incorrectly
VSTDIC001	VL042351.D	Cyclohexane	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	Ethanol	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	Heptane	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	m/p-Xylene	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	Methyl Methacrylate	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC001	VL042351.D	t-1,3-Dichloropropene	SAM	4/18/2025 4:19:17 PM	MMDado da	4/18/2025 10:45:15 PM	Peak Integrated by Software incorrectly
VSTDIC005	VL042352.D	1,1,2-Trichloroethane	SAM	4/18/2025 4:20:33 PM	MMDado da	4/18/2025 10:45:07 PM	Peak Integrated by Software

							e incorrect ly
VSTDIC0. 5	VL042352 .D	1,2- Dichloropropan e	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	1,4-Dioxane	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	4-Methyl-2- Pentanone	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	Chlorobenzene	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	cis-1,3- Dichloropropen e	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	Ethanol	SAM	4/18/20 25 4:20:33 PM	MMDado da	4/18/20 25 10:45:0 7 PM	Peak Integrat ed by Softwar e incorrect ly
VSTDIC0. 5	VL042352 .D	Heptane	SAM	4/18/20 25 4:20:33	MMDado da	4/18/20 25 10:45:0	Peak Integrat ed by

				PM		7 PM	Software incorrectly
VSTDIC0.5	VL042352.D	m/p-Xylene	SAM	4/18/2025 4:20:33 PM	MMDado da	4/18/2025 10:45:07 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042352.D	Methyl Methacrylate	SAM	4/18/2025 4:20:33 PM	MMDado da	4/18/2025 10:45:07 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042352.D	t-1,3-Dichloropropene	SAM	4/18/2025 4:20:33 PM	MMDado da	4/18/2025 10:45:07 PM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL042352.D	Tetrahydrofuran	SAM	4/18/2025 4:20:33 PM	MMDado da	4/18/2025 10:45:07 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042353.D	1,1,1-Trichloroethane	SAM	4/18/2025 4:20:37 PM	MMDado da	4/18/2025 10:45:04 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042353.D	1,1,2,2-Tetrachloroethane	SAM	4/18/2025 4:20:37 PM	MMDado da	4/18/2025 10:45:04 PM	Peak Integrated by Software incorrectly
VSTDIC0.1	VL042353.D	1,2-Dibromoethane	SAM	4/18/2025	MMDado da	4/18/2025	Peak Integrated

		e		4:20:37 PM		10:45:04 PM	ed by Software incorrectly
VSTDICC0.1	VL042353.D	Tetrachloroethene	SAM	4/18/2025 4:20:37 PM	MMDado da	4/18/2025 10:45:04 PM	Peak Integrated by Software incorrectly
VSTDICC0.1	VL042353.D	Trichloroethene	SAM	4/18/2025 4:20:37 PM	MMDado da	4/18/2025 10:45:04 PM	Peak Integrated by Software incorrectly
VSTDICC0.03	VL042354.D	1,1,1-Trichloroethane	SAM	4/18/2025 4:19:24 PM	MMDado da	4/18/2025 10:45:02 PM	Peak Integrated by Software incorrectly
VSTDICC0.03	VL042354.D	1,1,2,2-Tetrachloroethane	SAM	4/18/2025 4:19:24 PM	MMDado da	4/18/2025 10:45:02 PM	Peak Integrated by Software incorrectly
VSTDICC0.03	VL042354.D	Carbon Tetrachloride	SAM	4/18/2025 4:19:24 PM	MMDado da	4/18/2025 10:45:02 PM	Peak Integrated by Software incorrectly
VSTDICC0.03	VL042354.D	Tetrachloroethene	SAM	4/18/2025 4:19:24 PM	MMDado da	4/18/2025 10:45:02 PM	Peak Integrated by Software incorrectly
VSTDICC0.	VL042354	Trichloroethene	SAM	4/18/20	MMDado	4/18/20	Peak

03	.D	e		25 4:19:24 PM	da	25 10:45:02 PM	Integrat ed by Softwar e incorrec tly
VSTDIC015	VL042355.D	m/p-Xylene	SAM	4/18/2025 4:21:35 PM	MMDado da	4/18/2025 10:44:59 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDICV010	VL042356.D	m/p-Xylene	SAM	4/18/2025 4:20:42 PM	MMDado da	4/18/2025 10:44:57 PM	Peak Integrat ed by Softwar e incorrec tly

Manual Integration Report			
Sequence	VL042825	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL042416.D	m/p-Xylene	SAM	4/29/2025 9:26:25 AM	MMDado da	4/30/2025 2:28:57 PM	Peak Integrat ed by Softwar e incorrec tly
VL0428ABS01	VL042418.D	m/p-Xylene	SAM	4/29/2025 9:26:31 AM	MMDado da	4/30/2025 2:29:13 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-02	VL042426.D	Benzene	SAM	4/29/2025 9:29:26 AM	MMDado da	4/30/2025 2:29:45 PM	Peak Integrat ed by Softwar e incorrec

							tly
Q1880-02	VL04242 6.D	Carbon Tetrachloride	SAM	4/29/20 25 9:29:26 AM	MMDado da	4/30/20 25 2:29:45 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-02	VL04242 6.D	Chlorodifluoromet hane	SAM	4/29/20 25 9:29:26 AM	MMDado da	4/30/20 25 2:29:45 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-02	VL04242 6.D	m/p-Xylene	SAM	4/29/20 25 9:29:26 AM	MMDado da	4/30/20 25 2:29:45 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-03	VL04242 7.D	Carbon Tetrachloride	SAM	4/29/20 25 9:28:00 AM	MMDado da	4/30/20 25 2:29:48 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-03	VL04242 7.D	Chlorodifluoromet hane	SAM	4/29/20 25 9:28:00 AM	MMDado da	4/30/20 25 2:29:48 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-03	VL04242 7.D	Heptane	SAM	4/29/20 25 9:28:00 AM	MMDado da	4/30/20 25 2:29:48 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-03	VL04242 7.D	Propene	SAM	4/29/20 25 9:28:00 AM	MMDado da	4/30/20 25 2:29:48 PM	Peak Integrat ed by Softwar e

							incorrectly
Q1880-03	VL04242 7.D	Toluene	SAM	4/29/20 25 9:28:00 AM	MMDado da	4/30/20 25 2:29:48 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-08	VL04242 8.D	Acetone	SAM	4/29/20 25 9:26:42 AM	MMDado da	4/30/20 25 2:29:49 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-08	VL04242 8.D	Carbon Tetrachloride	SAM	4/29/20 25 9:26:42 AM	MMDado da	4/30/20 25 2:29:49 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-08	VL04242 8.D	Chlorodifluoromet hane	SAM	4/29/20 25 9:26:42 AM	MMDado da	4/30/20 25 2:29:49 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-08	VL04242 8.D	Cyclohexane	SAM	4/29/20 25 9:26:42 AM	MMDado da	4/30/20 25 2:29:49 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-08	VL04242 8.D	m/p-Xylene	SAM	4/29/20 25 9:26:42 AM	MMDado da	4/30/20 25 2:29:49 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-10	VL04242 9.D	Carbon Tetrachloride	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar

							e incorec tly
Q1880-10	VL04242 9.D	Chlorodifluoromet hane	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-10	VL04242 9.D	m/p-Xylene	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-10	VL04242 9.D	Propene	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-10	VL04242 9.D	Tetrachloroethene	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-10	VL04242 9.D	Toluene	SAM	4/29/20 25 9:28:05 AM	MMDado da	4/30/20 25 2:29:52 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880- 10DUP	VL04243 0.D	Benzene	SAM	4/29/20 25 9:28:10 AM	MMDado da	4/30/20 25 2:29:59 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880- 10DUP	VL04243 0.D	Carbon Tetrachloride	SAM	4/29/20 25 9:28:10	MMDado da	4/30/20 25 2:29:59	Peak Integrat ed by

				AM		PM	Software incorrectly
Q1880-10DUP	VL04243 0.D	Chlorodifluoromethane	SAM	4/29/2025 9:28:10 AM	MMDado da	4/30/2025 2:29:59 PM	Peak Integrated by Software incorrectly
Q1880-10DUP	VL04243 0.D	m/p-Xylene	SAM	4/29/2025 9:28:10 AM	MMDado da	4/30/2025 2:29:59 PM	Peak Integrated by Software incorrectly
Q1880-10DUP	VL04243 0.D	Tetrachloroethene	SAM	4/29/2025 9:28:10 AM	MMDado da	4/30/2025 2:29:59 PM	Peak Integrated by Software incorrectly
Q1880-10DUP	VL04243 0.D	Toluene	SAM	4/29/2025 9:28:10 AM	MMDado da	4/30/2025 2:29:59 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	4-Methyl-2-Pentanone	SAM	4/29/2025 9:27:16 AM	MMDado da	4/30/2025 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	Carbon Tetrachloride	SAM	4/29/2025 9:27:16 AM	MMDado da	4/30/2025 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	Chlorodifluoromethane	SAM	4/29/2025	MMDado da	4/30/2025	Peak Integrated

				9:27:16 AM		2:30:37 PM	ed by Software incorrectly
Q1880-01	VL04243 1.D	Cyclohexane	SAM	4/29/20 25 9:27:16 AM	MMDado da	4/30/20 25 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	Heptane	SAM	4/29/20 25 9:27:16 AM	MMDado da	4/30/20 25 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	o-Xylene	SAM	4/29/20 25 9:27:16 AM	MMDado da	4/30/20 25 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-01	VL04243 1.D	Trichloroethene	SAM	4/29/20 25 9:27:16 AM	MMDado da	4/30/20 25 2:30:37 PM	Peak Integrated by Software incorrectly
Q1880-06	VL04243 2.D	Carbon Tetrachloride	SAM	4/29/20 25 9:28:14 AM	MMDado da	4/30/20 25 2:30:01 PM	Peak Integrated by Software incorrectly
Q1880-06	VL04243 2.D	Chlorodifluoromethane	SAM	4/29/20 25 9:28:14 AM	MMDado da	4/30/20 25 2:30:01 PM	Peak Integrated by Software incorrectly
Q1880-06	VL04243	Toluene	SAM	4/29/20	MMDado	4/30/20	Peak

	2.D			25 9:28:14 AM	da	25 2:30:01 PM	Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	1,2,4- Trimethylbenzene	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	Carbon Tetrachloride	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	Chlorodifluoromet hane	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	m/p-Xylene	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	Tetrachloroethene	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-09	VL04243 3.D	Toluene	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly

Q1880-09	VL04243 3.D	Trichlorofluoromet hane	SAM	4/29/20 25 9:27:21 AM	MMDado da	4/30/20 25 2:30:03 PM	Peak Integrat ed by Softwar e incorrec tly
----------	----------------	----------------------------	-----	--------------------------------	--------------	--------------------------------	---

Manual Integration Report			
Sequence	VL042925	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL04243 6.D	1,1,2- Trichloroethane	SAM	4/30/20 25 7:49:08 AM	MMDado da	4/30/20 25 2:26:33 PM	Peak Integrat ed by Softwar e incorrec tly
VSTDCCC010	VL04243 6.D	m/p-Xylene	SAM	4/30/20 25 7:49:08 AM	MMDado da	4/30/20 25 2:26:33 PM	Peak Integrat ed by Softwar e incorrec tly
VL0429AB S01	VL04243 8.D	m/p-Xylene	SAM	4/30/20 25 7:49:14 AM	MMDado da	4/30/20 25 2:26:38 PM	Peak Integrat ed by Softwar e incorrec tly
VL0429AB S01	VL04243 8.D	Methyl Methacrylate	SAM	4/30/20 25 7:49:14 AM	MMDado da	4/30/20 25 2:26:38 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-04	VL04243 9.D	Carbon Tetrachloride	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrat ed by Softwar e

							incorrectly
Q1880-04	VL04243 9.D	Chlorodifluoromethane	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-04	VL04243 9.D	Hexane	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-04	VL04243 9.D	m/p-Xylene	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-04	VL04243 9.D	Tetrachloroethene	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-04	VL04243 9.D	Toluene	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-04	VL04243 9.D	Trichlorofluoromethane	SAM	4/30/20 25 7:50:22 AM	MMDado da	4/30/20 25 2:26:43 PM	Peak Integrated by Software incorrectly
Q1880-07DL	VL04244 1.D	2-Butanone	SAM	4/30/20 25 7:51:22 AM	MMDado da	4/30/20 25 2:28:23 PM	Peak Integrated by Software

							e in cor rec tly
Q1880-05	VL04244 7.D	Chlorodifluorometh ane	SAM	4/30/20 25 7:53:11 AM	MMDado da	4/30/20 25 2:27:30 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-05	VL04244 7.D	m/p-Xylene	SAM	4/30/20 25 7:53:11 AM	MMDado da	4/30/20 25 2:27:30 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-05	VL04244 7.D	Tetrachloroethene	SAM	4/30/20 25 7:53:11 AM	MMDado da	4/30/20 25 2:27:30 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-05	VL04244 7.D	Toluene	SAM	4/30/20 25 7:53:11 AM	MMDado da	4/30/20 25 2:27:30 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-05	VL04244 7.D	Trichlorofluoromet hane	SAM	4/30/20 25 7:53:11 AM	MMDado da	4/30/20 25 2:27:30 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-07	VL04244 8.D	Benzene	SAM	4/30/20 25 7:52:30 AM	MMDado da	4/30/20 25 2:27:33 PM	Peak Integrat ed by Softwar e in cor rec tly
Q1880-07	VL04244 8.D	Carbon Tetrachloride	SAM	4/30/20 25 7:52:30	MMDado da	4/30/20 25 2:27:33	Peak Integrat ed by

				AM		PM	Software incorrectly
Q1880-07	VL04244 8.D	Chlorodifluoromethane	SAM	4/30/2025 7:52:30 AM	MMDado da	4/30/2025 2:27:33 PM	Peak Integrated by Software incorrectly
Q1880-07	VL04244 8.D	Dichlorodifluoromethane	SAM	4/30/2025 7:52:30 AM	MMDado da	4/30/2025 2:27:33 PM	Peak Integrated by Software incorrectly
Q1880-07	VL04244 8.D	m/p-Xylene	SAM	4/30/2025 7:52:30 AM	MMDado da	4/30/2025 2:27:33 PM	Peak Integrated by Software incorrectly
Q1880-07	VL04244 8.D	Tetrahydrofuran	SAM	4/30/2025 7:52:30 AM	MMDado da	4/30/2025 2:27:33 PM	Peak Integrated by Software incorrectly
Q1880-07	VL04244 8.D	Toluene	SAM	4/30/2025 7:52:30 AM	MMDado da	4/30/2025 2:27:33 PM	Peak Integrated by Software incorrectly
Q1880-11	VL04244 9.D	Carbon Tetrachloride	SAM	4/30/2025 7:49:52 AM	MMDado da	4/30/2025 2:27:40 PM	Peak Integrated by Software incorrectly
Q1880-11	VL04244 9.D	Chlorodifluoromethane	SAM	4/30/2025	MMDado da	4/30/2025	Peak Integrated

				7:49:52 AM		2:27:40 PM	ed by Softwar e incorrec tly
Q1880-11	VL04244 9.D	Hexane	SAM	4/30/20 25 7:49:52 AM	MMDado da	4/30/20 25 2:27:40 PM	Peak Integrat ed by Softwar e incorrec tly
Q1880-11	VL04244 9.D	Toluene	SAM	4/30/20 25 7:49:52 AM	MMDado da	4/30/20 25 2:27:40 PM	Peak Integrat ed by Softwar e incorrec tly

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____