

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR16

Lab Code: CHEM Case No.: Q1880 SAS No.: Q1880 SDG No.: Q1880

Instrument ID: MSVOA_L Calibration Date/Time: 04/29/2025 08:39

Lab File ID: VL042436.D Init. Calib. Date(s): 04/17/2025 04/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:44 12:54

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.767	2.050		16.02	30
Chloromethane	0.699	0.788		12.73	30
Vinyl Chloride	0.761	0.756		-0.66	30
Bromomethane	0.327	0.358		9.48	30
Chloroethane	0.275	0.307		11.64	30
Tetrahydrofuran	0.461	0.391		-15.18	30
Trichlorofluoromethane	1.680	1.990		18.45	30
1,1,2-Trichlorotrifluoroethane	1.374	1.502		9.32	30
Dichlorotetrafluoroethane	1.517	1.746		15.1	30
tert-Butyl alcohol	2.029	2.005		-1.18	30
Heptane	2.153	1.613		-25.08	30
1,1-Dichloroethene	0.642	0.659		2.65	30
Acetone	1.637	1.614		-1.4	30
Carbon Disulfide	1.749	1.858		6.23	30
Methyl tert-Butyl Ether	1.070	0.996		-6.92	30
Methylene Chloride	0.641	0.587		-8.42	30
trans-1,2-Dichloroethene	0.731	0.756		3.42	30
1,1-Dichloroethane	1.388	1.506		8.5	30
Cyclohexane	1.320	1.140		-13.64	30
2-Butanone	0.791	0.745		-5.82	30
Carbon Tetrachloride	0.905	1.017		12.38	30
cis-1,2-Dichloroethene	1.553	1.525		-1.8	30
Chloroform	2.309	2.540		10	30
1,1,1-Trichloroethane	2.599	2.727		4.93	30
2,2,4-Trimethylpentane	1.863	1.678		-9.93	30
Benzene	1.079	1.010		-6.39	30
1,2-Dichloroethane	0.647	0.786		21.48	30
Trichloroethene	0.471	0.398		-15.5	30
1,2-Dichloropropane	0.402	0.354		-11.94	30
Bromodichloromethane	0.901	1.002		11.21	30
4-Methyl-2-Pentanone	1.314	1.047		-20.32	30
Toluene	1.294	1.161		-10.28	30
t-1,3-Dichloropropene	0.497	0.448		-9.86	30
cis-1,3-Dichloropropene	0.608	0.544		-10.53	30
1,1,2-Trichloroethane	0.410	0.415		1.22	30
Dibromochloromethane	0.723	0.748		3.46	30
1,2-Dibromoethane	0.586	0.578		-1.37	30
Tetrachloroethene	0.417	0.327		-21.58	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR16

Lab Code: CHEM Case No.: Q1880 SAS No.: Q1880 SDG No.: Q1880

Instrument ID: MSVOA_L Calibration Date/Time: 04/29/2025 08:39

Lab File ID: VL042436.D Init. Calib. Date(s): 04/17/2025 04/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:44 12:54

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	1.040	1.049		0.87	30
Ethyl Benzene	1.981	1.988		0.35	30
m/p-Xylene	1.596	1.685		5.58	30
o-Xylene	1.607	1.718		6.91	30
Styrene	0.767	0.725		-5.48	30
Bromoform	0.633	0.607		-4.11	30
1,1,2,2-Tetrachloroethane	1.042	1.010		-3.07	30
2-Chlorotoluene	1.851	1.956		5.67	30
1,3,5-Trimethylbenzene	1.735	1.830		5.48	30
1,2,4-Trimethylbenzene	2.026	2.052		1.28	30
1,3-Dichlorobenzene	1.053	1.056		0.28	30
1,4-Dichlorobenzene	1.059	1.081		2.08	30
1,2-Dichlorobenzene	1.024	1.051		2.64	30
1,2,4-Trichlorobenzene	1.040	0.974		-6.35	30
Hexachloro-1,3-Butadiene	1.034	0.976		-5.61	30
1,3-Butadiene	0.783	0.776		-0.89	30
Naphthalene	2.423	2.279		-5.94	30
4-Ethyltoluene	1.965	2.072		5.45	30
1-Bromo-4-Fluorobenzene	0.822	0.938		14.11	30
Hexane	1.633	1.343		-17.76	30
Allyl Chloride	1.130	1.151		1.86	30
1,4-Dioxane	220.893	174.653		-20.93	30
Methyl Methacrylate	0.486	0.407		-16.25	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.