

LB135594


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

69617



A= 0.0000e+000
 B= 1.4399e-004
 C= 4.9983e-002
 Rho= 0.9996920
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%AD
0.0	0.000	0.063	0.063	92	0.000	92	0				19
0.2	0.200	0.238	0.038	1308	0.0 %	1308					6
2.5	2.500	2.357	-0.143	16019	0.0 %	16019					2
5.0	5.000	5.080	0.080	34930	0.0 %	34930					1
7.5	7.500	7.388	-0.112	50965	0.0 %	50965					1
10.0	10.000	10.074	0.074	69617	0.0 %	69617					

LB135594

Method: 7470A

Operator: Admin

Date of Analysis: 29 Apr 2025 13:14:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	92	-	0.0000	7470A	PPB	29 Apr 2025 13:16:09	S	Std
0.2 - 1	51.2	1308	-	0.2000	7470A	PPB	29 Apr 2025 13:21:03	S	Std
2.5 - 1	52.5	16019	-	2.5000	7470A	PPB	29 Apr 2025 13:23:20	S	Std
5.0 - 1	53	34930	-	5.0000	7470A	PPB	29 Apr 2025 13:25:37	S	Std
7.5 - 1	54.5	50965	-	7.5000	7470A	PPB	29 Apr 2025 13:33:22	S	Std
10.0 - 1	510	69617	-	10.0000	7470A	PPB	29 Apr 2025 13:46:21	S	Std
ICV106 - 1	ICV106	26209	3.8238	-	7470A	PPB	29 Apr 2025 13:51:49	U	SMPL
ICB106 - 1	ICB106	-314	0.0048	-	7470A	PPB	29 Apr 2025 13:54:04	U	SMPL
CCV43 - 1	CCV43	34863	5.0699	-	7470A	PPB	29 Apr 2025 13:56:22	U	SMPL
CCB43 - 1	CCB43	-376	-0.0042	-	7470A	PPB	29 Apr 2025 13:58:37	U	SMPL
CRA - 1	CRA	1203	0.2232	-	7470A	PPB	29 Apr 2025 14:00:54	U	SMPL
HighStd - 1	HighStd	69034	9.9901	-	7470A	PPB	29 Apr 2025 14:03:09	U	SMPL
ChkStd - 1	ChkStd	43171	6.2661	-	7470A	PPB	29 Apr 2025 14:05:25	U	SMPL
PB167789BL - 1	PBW	-525	-0.0256	-	7470A	PPB	29 Apr 2025 14:07:48	U	SMPL
PB167789BS - 1	LCSW	30967	4.5089	-	7470A	PPB	29 Apr 2025 14:12:40	U	SMPL
Q1883-02 - 1	OU4-PCS-TC-27-042325	1619	0.2831	-	7470A	PPB	29 Apr 2025 14:14:57	U	SMPL
Q1883-04 - 1	OU4-PCS-TC-28-042325	-98	0.0359	-	7470A	PPB	29 Apr 2025 14:17:15	U	SMPL
Q1883-06 - 1	OU4-PCS-TC-29-042325	5	0.0507	-	7470A	PPB	29 Apr 2025 14:19:35	U	SMPL
Q1883-08 - 1	OU4-PCS-TC-30-042325	28	0.0540	-	7470A	PPB	29 Apr 2025 14:21:54	U	SMPL
Q1883-10 - 1	OU4-PCS-TC-31-042325	53	0.0576	-	7470A	PPB	29 Apr 2025 14:24:09	U	SMPL
CCV44 - 1	CCV44	33748	4.9093	-	7470A	PPB	29 Apr 2025 14:26:24	U	SMPL
CCB44 - 1	CCB44	-229	0.0170	-	7470A	PPB	29 Apr 2025 14:28:39	U	SMPL
Q1883-12 - 1	OU4-PCS-TC-32-042325	250	0.0860	-	7470A	PPB	29 Apr 2025 14:30:57	U	SMPL
Q1883-14 - 1	OU4-VSL-18-042325	8	0.0511	-	7470A	PPB	29 Apr 2025 14:33:12	U	SMPL
Q1883-16 - 1	OU4-VSL-19-042325	37	0.0553	-	7470A	PPB	29 Apr 2025 14:35:28	U	SMPL
Q1883-16DUP - 1	OU4-VSL-19-042325DUP	59	0.0585	-	7470A	PPB	29 Apr 2025 14:37:44	U	SMPL
Q1883-16MS - 1	OU4-VSL-19-042325MS	18318	2.6876	-	7470A	PPB	29 Apr 2025 14:39:59	U	SMPL
Q1883-16MSD - 1	OU4-VSL-19-042325MSD	17810	2.6144	-	7470A	PPB	29 Apr 2025 14:42:16	U	SMPL
PB167774TB - 1	PB167774TB	45	0.0565	-	7470A	PPB	29 Apr 2025 14:46:39	U	SMPL
Q1883-16LX5 - 1	PB167789TB	48	0.0569	-	7470A	PPB	29 Apr 2025 14:48:56	U	SMPL
Q1883-16A - 1		27767	4.0481	-	7470A	PPB	29 Apr 2025 14:51:13	U	SMPL
CCV45 - 1	CCV45	33701	4.9025	-	7470A	PPB	29 Apr 2025 14:53:28	U	SMPL
CCB45 - 1	CCB45	-249	0.0141	-	7470A	PPB	29 Apr 2025 14:55:45	U	SMPL

MB

05/07/25