

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135594

Review By	mohan	Review On	5/1/2025 9:41:29 AM
Supervise By	jaswal	Supervise On	5/1/2025 11:29:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP85479,MP85480,MP85481,MP85482,MP85483,MP85484
ICV Standard	MP85485
CCV Standard	MP85487
ICSA Standard	
CRI Standard	MP85489
LCS Standard	
Chk Standard	MP85486,MP85488,MP85490,MP85492

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	04/29/25 13:16		mohan	OK
2	S0.2	S0.2	CAL2	04/29/25 13:21		mohan	OK
3	S2.5	S2.5	CAL3	04/29/25 13:23		mohan	OK
4	S5	S5	CAL4	04/29/25 13:25		mohan	OK
5	S7.5	S7.5	CAL5	04/29/25 13:33		mohan	OK
6	S10	S10	CAL6	04/29/25 13:46		mohan	OK
7	ICV106	ICV106	ICV	04/29/25 13:51		mohan	OK
8	ICB106	ICB106	ICB	04/29/25 13:54		mohan	OK
9	CCV43	CCV43	CCV	04/29/25 13:56		mohan	OK
10	CCB43	CCB43	CCB	04/29/25 13:58		mohan	OK
11	CRA	CRA	CRDL	04/29/25 14:00		mohan	OK
12	HighStd	HighStd	HIGH STD	04/29/25 14:03		mohan	OK
13	ChkStd	ChkStd	SAM	04/29/25 14:05		mohan	OK
14	PB167789BL	PB167789BL	MB	04/29/25 14:07		mohan	OK
15	PB167789BS	PB167789BS	LCS	04/29/25 14:12		mohan	OK
16	Q1883-02	OU4-PCS-TC-27-042	SAM	04/29/25 14:14		mohan	OK
17	Q1883-04	OU4-PCS-TC-28-042	SAM	04/29/25 14:17		mohan	OK
18	Q1883-06	OU4-PCS-TC-29-042	SAM	04/29/25 14:19		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135594

Review By	mohan	Review On	5/1/2025 9:41:29 AM
Supervise By	jaswal	Supervise On	5/1/2025 11:29:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP85479,MP85480,MP85481,MP85482,MP85483,MP85484
ICV Standard	MP85485
CCV Standard	MP85487
ICSA Standard	
CRI Standard	MP85489
LCS Standard	
Chk Standard	MP85486,MP85488,MP85490,MP85492

19	Q1883-08	OU4-PCS-TC-30-042	SAM	04/29/25 14:21		mohan	OK
20	Q1883-10	OU4-PCS-TC-31-042	SAM	04/29/25 14:24		mohan	OK
21	CCV44	CCV44	CCV	04/29/25 14:26		mohan	OK
22	CCB44	CCB44	CCB	04/29/25 14:28		mohan	OK
23	Q1883-12	OU4-PCS-TC-32-042	SAM	04/29/25 14:30		mohan	OK
24	Q1883-14	OU4-VSL-18-042325	SAM	04/29/25 14:33		mohan	OK
25	Q1883-16	OU4-VSL-19-042325	SAM	04/29/25 14:35		mohan	OK
26	Q1883-16DUP	OU4-VSL-19-042325	DUP	04/29/25 14:37		mohan	OK
27	Q1883-16MS	OU4-VSL-19-042325	MS	04/29/25 14:39		mohan	OK
28	Q1883-16MSD	OU4-VSL-19-042325	MSD	04/29/25 14:42		mohan	OK
29	PB167789TB	PB167789TB	MB	04/29/25 14:46		mohan	OK
30	Q1883-16L	OU4-VSL-19-042325	SD	04/29/25 14:48		mohan	OK
31	Q1883-16A	OU4-VSL-19-042325	PS	04/29/25 14:51		mohan	OK
32	CCV45	CCV45	CCV	04/29/25 14:53		mohan	OK
33	CCB45	CCB45	CCB	04/29/25 14:55		mohan	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB135595

Review By	jaswal	Review On	4/29/2025 4:08:50 PM
Supervise By	mohan	Supervise On	4/29/2025 4:10:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP85351,MP85362,MP85361,MP85358,MP85357,MP85356,MP85355,MP85353,MP85352
ICV Standard	MP85363,MP85361
CCV Standard	MP85365
ICSA Standard	MP85366,MP85367
CRI Standard	MP85362
LCS Standard	
Chk Standard	MP85369,MP85370

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	04/29/25 10:50		Jaswal	OK
2	S0	S0	CAL1	04/29/25 11:31		Jaswal	OK
3	S2	S2	CAL3	04/29/25 11:38		Jaswal	OK
4	S3	S3	CAL4	04/29/25 11:41		Jaswal	OK
5	S4	S4	CAL5	04/29/25 11:45		Jaswal	OK
6	S5	S5	CAL6	04/29/25 11:47		Jaswal	OK
7	S6	S6	CAL7	04/29/25 11:50		Jaswal	OK
8	S7	S7	CAL8	04/29/25 11:53		Jaswal	OK
9	S8	S8	CAL9	04/29/25 11:56		Jaswal	OK
10	ICV01	ICV01	ICV	04/29/25 12:14		Jaswal	OK
11	LLICV	LLICV	LLICV	04/29/25 12:18		Jaswal	OK
12	ICB01	ICB01	ICB	04/29/25 12:29		Jaswal	OK
13	ICSA01	ICSA01	ICSA	04/29/25 12:32		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	04/29/25 12:37		Jaswal	OK
15	CCV01	CCV01	CCV	04/29/25 12:40		Jaswal	OK
16	CCB01	CCB01	CCB	04/29/25 12:44		Jaswal	OK
17	CRI	CRI	CRDL	04/29/25 12:47		Jaswal	OK
18	Q1812-02	RINSE-EB-TANK-041	SAM	04/29/25 12:50		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB135595

Review By	jaswal	Review On	4/29/2025 4:08:50 PM
Supervise By	mohan	Supervise On	4/29/2025 4:10:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85351,MP85362,MP85361,MP85358,MP85357,MP85356,MP85355,MP85353,MP85352		
ICV Standard	MP85363,MP85361		
CCV Standard	MP85365		
ICSA Standard	MP85366,MP85367		
CRI Standard	MP85362		
LCS Standard			
Chk Standard	MP85369,MP85370		

19	Q1812-03	RINSE-EB-PUMP-04	SAM	04/29/25 12:54		Jaswal	OK
20	PB167612BL	PB167612BL	MB	04/29/25 12:57		Jaswal	OK
21	PB167612BS	PB167612BS	LCS	04/29/25 13:00		Jaswal	OK
22	CCV02	CCV02	CCV	04/29/25 13:03		Jaswal	OK
23	CCB02	CCB02	CCB	04/29/25 13:07		Jaswal	OK
24	CCV03	CCV03	CCV	04/29/25 14:47		Jaswal	OK
25	CCB03	CCB03	CCB	04/29/25 14:59		Jaswal	OK
26	PB167780BL	PB167780BL	MB	04/29/25 15:04		Jaswal	OK
27	PB167780BS	PB167780BS	LCS	04/29/25 15:08		Jaswal	OK
28	PB167780TB	PB167780TB	MB	04/29/25 15:15		Jaswal	OK
29	Q1883-02DL	OU4-PCS-TC-27-042	SAM	04/29/25 15:18	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
30	Q1883-04DL	OU4-PCS-TC-28-042	SAM	04/29/25 15:21	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
31	Q1883-06DL	OU4-PCS-TC-29-042	SAM	04/29/25 15:25	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
32	Q1883-08DL	OU4-PCS-TC-30-042	SAM	04/29/25 15:28	INT_STD fail for 89Y (1)	Jaswal	Dilution
33	Q1883-10DL	OU4-PCS-TC-31-042	SAM	04/29/25 15:31	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
34	Q1883-12DL	OU4-PCS-TC-32-042	SAM	04/29/25 15:34	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
35	CCV04	CCV04	CCV	04/29/25 15:41		Jaswal	OK
36	CCB04	CCB04	CCB	04/29/25 15:48		Jaswal	OK
37	Q1883-14DL	OU4-VSL-18-042325	SAM	04/29/25 15:55	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
38	Q1883-16DL	OU4-VSL-19-042325	SAM	04/29/25 15:58	INT_STD fail for 89Y (1,2)	Jaswal	Dilution

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB135595

Review By	jaswal	Review On	4/29/2025 4:08:50 PM
Supervise By	mohan	Supervise On	4/29/2025 4:10:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP85351,MP85362,MP85361,MP85358,MP85357,MP85356,MP85355,MP85353,MP85352
ICV Standard	MP85363,MP85361
CCV Standard	MP85365
ICSA Standard	MP85366,MP85367
CRI Standard	MP85362
LCS Standard	
Chk Standard	MP85369,MP85370

39	Q1883-16DUPDL	OU4-VSL-19-042325	DUP	04/29/25 16:01	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
40	Q1883-16LDL	OU4-VSL-19-042325	SD	04/29/25 16:04	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
41	Q1883-16MSDL	OU4-VSL-19-042325	MS	04/29/25 16:08	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
42	Q1883-16MSDDL	OU4-VSL-19-042325	MSD	04/29/25 16:11	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
43	Q1883-16ADL	OU4-VSL-19-042325	PS	04/29/25 16:14	INT_STD fail for 89Y (1,2)	Jaswal	Dilution
44	CCV05	CCV05	CCV	04/29/25 16:20		Jaswal	OK
45	CCB05	CCB05	CCB	04/29/25 16:28		Jaswal	OK
46	Q1883-02DL2	OU4-PCS-TC-27-0423	SAM	04/29/25 16:36	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
47	Q1883-04DL2	OU4-PCS-TC-28-0423	SAM	04/29/25 16:40	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
48	Q1883-06DL2	OU4-PCS-TC-29-0423	SAM	04/29/25 16:43	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
49	Q1883-08DL2	OU4-PCS-TC-30-0423	SAM	04/29/25 16:46	25x for INT_STD fail for 89Y (1)	Jaswal	Confirms
50	Q1883-10DL2	OU4-PCS-TC-31-0423	SAM	04/29/25 16:50	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
51	Q1883-12DL2	OU4-PCS-TC-32-0423	SAM	04/29/25 16:53	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
52	Q1883-14DL2	OU4-VSL-18-042325	SAM	04/29/25 16:56	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
53	CCV06	CCV06	CCV	04/29/25 17:03		Jaswal	OK
54	CCB06	CCB06	CCB	04/29/25 17:05		Jaswal	OK
55	Q1883-16DL2	OU4-VSL-19-042325	SAM	04/29/25 17:09	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB135595

Review By	jaswal	Review On	4/29/2025 4:08:50 PM
Supervise By	mohan	Supervise On	4/29/2025 4:10:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85351,MP85362,MP85361,MP85358,MP85357,MP85356,MP85355,MP85353,MP85352		
ICV Standard	MP85363,MP85361		
CCV Standard	MP85365		
ICSA Standard	MP85366,MP85367		
CRI Standard	MP85362		
LCS Standard			
Chk Standard	MP85369,MP85370		

56	Q1883-16DUPDL2	OU4-VSL-19-042325	DUP	04/29/25 17:12	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
57	Q1883-16LDL2	OU4-VSL-19-042325	SD	04/29/25 17:15	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
58	Q1883-16MSDL2	OU4-VSL-19-042325	MS	04/29/25 17:19	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
59	Q1883-16MSDDL2	OU4-VSL-19-042325	MSD	04/29/25 17:22	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
60	Q1883-16ADL2	OU4-VSL-19-042325	PS	04/29/25 17:25	25x for INT_STD fail for 89Y (1,2)	Jaswal	Confirms
61	CCV07	CCV07	CCV	04/29/25 17:41		Jaswal	OK
62	CCB07	CCB07	CCB	04/29/25 17:50		Jaswal	OK