



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b>   | <u>Nobis Group</u> |                  |                     |                  | <b>SDG No.:</b> | <u>Q1883</u>    |              |                  |                  |               |
|------------------|--------------------|------------------|---------------------|------------------|-----------------|-----------------|--------------|------------------|------------------|---------------|
| <b>Contract:</b> | <u>NOBI03</u>      | <b>Lab Code:</b> | <u>CHEM</u>         | <b>Case No.:</b> | <u>Q1883</u>    | <b>SAS No.:</b> | <u>Q1883</u> |                  |                  |               |
| Sample ID        | Analyte            | Result<br>ug/L   | Acceptance<br>Limit | Conc<br>Qual     | LOD             | CRQL            | M            | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| ICB106           | Mercury            | 0.20             | +/-0.20             | U                | 0.16            | 0.20            | CV           | 04/29/2025       | 13:54            | LB135594      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b>   | <u>Nobis Group</u> |                  |                     |                  | <b>SDG No.:</b> | <u>Q1883</u>    |              |                  |                  |               |
|------------------|--------------------|------------------|---------------------|------------------|-----------------|-----------------|--------------|------------------|------------------|---------------|
| <b>Contract:</b> | <u>NOBI03</u>      | <b>Lab Code:</b> | <u>CHEM</u>         | <b>Case No.:</b> | <u>Q1883</u>    | <b>SAS No.:</b> | <u>Q1883</u> |                  |                  |               |
| Sample ID        | Analyte            | Result<br>ug/L   | Acceptance<br>Limit | Conc<br>Qual     | LOD             | CRQL            | M            | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB43            | Mercury            | 0.20             | +/-0.20             | U                | 0.16            | 0.20            | CV           | 04/29/2025       | 13:58            | LB135594      |
| CCB44            | Mercury            | 0.20             | +/-0.20             | U                | 0.16            | 0.20            | CV           | 04/29/2025       | 14:28            | LB135594      |
| CCB45            | Mercury            | 0.20             | +/-0.20             | U                | 0.16            | 0.20            | CV           | 04/29/2025       | 14:55            | LB135594      |

# Metals

- 3a -

## INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Nobis Group **SDG No.:** Q1883  
**Contract:** NOBI03 **Lab Code:** CHEM **Case No.:** Q1883 **SAS No.:** Q1883

| Sample ID | Analyte   | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|-----------|----------------|---------------------|--------------|------|------|---|------------------|------------------|---------------|
| ICB01     | Antimony  | 2.00           | +/-2.00             | U            | 0.25 | 2.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Arsenic   | 1.00           | +/-1.00             | U            | 0.25 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Barium    | 10.0           | +/-10.0             | U            | 1.25 | 10.0 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Beryllium | 1.00           | +/-1.00             | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Cadmium   | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Chromium  | 2.00           | +/-2.00             | U            | 0.75 | 2.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Copper    | 2.00           | +/-2.00             | U            | 1.50 | 2.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Lead      | 1.00           | +/-1.00             | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Nickel    | 1.00           | +/-1.00             | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Selenium  | 5.00           | +/-5.00             | U            | 4.50 | 5.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Silver    | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Thallium  | 1.00           | +/-1.00             | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Vanadium  | 5.00           | +/-5.00             | U            | 0.25 | 5.00 | P | 04/29/2025       | 12:29            | LB135595      |
|           | Zinc      | 5.00           | +/-5.00             | U            | 1.50 | 5.00 | P | 04/29/2025       | 12:29            | LB135595      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> <u>Nobis Group</u> |                              | <b>SDG No.:</b> <u>Q1883</u>  |                              |              |      |      |   |                  |                  |               |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|--------------|------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> <u>NOBI03</u>    | <b>Lab Code:</b> <u>CHEM</u> | <b>Case No.:</b> <u>Q1883</u> | <b>SAS No.:</b> <u>Q1883</u> |              |      |      |   |                  |                  |               |
| Sample ID                         | Analyte                      | Result<br>ug/L                | Acceptance<br>Limit          | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| <b>CCB01</b>                      | Antimony                     | 2.00                          | +/-2.00                      | U            | 0.25 | 2.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Lead                         | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Thallium                     | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Vanadium                     | 5.00                          | +/-5.00                      | U            | 0.25 | 5.00 | P | 04/29/2025       | 12:44            | LB135595      |
|                                   | Zinc                         | 5.00                          | +/-5.00                      | U            | 1.50 | 5.00 | P | 04/29/2025       | 12:44            | LB135595      |
| <b>CCB02</b>                      | Antimony                     | 2.00                          | +/-2.00                      | U            | 0.25 | 2.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Lead                         | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Thallium                     | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Vanadium                     | 5.00                          | +/-5.00                      | U            | 0.25 | 5.00 | P | 04/29/2025       | 13:07            | LB135595      |
|                                   | Zinc                         | 5.00                          | +/-5.00                      | U            | 1.50 | 5.00 | P | 04/29/2025       | 13:07            | LB135595      |
| <b>CCB03</b>                      | Antimony                     | 2.00                          | +/-2.00                      | U            | 0.25 | 2.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Lead                         | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 04/29/2025       | 14:59            | LB135595      |
|                                   | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 14:59            | LB135595      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> Nobis Group |           | <b>SDG No.:</b> Q1883 |                     |                  |       |                       |   |                  |                  |               |
|----------------------------|-----------|-----------------------|---------------------|------------------|-------|-----------------------|---|------------------|------------------|---------------|
| <b>Contract:</b>           | NOBI03    | <b>Lab Code:</b>      | CHEM                | <b>Case No.:</b> | Q1883 | <b>SAS No.:</b> Q1883 |   |                  |                  |               |
| Sample ID                  | Analyte   | Result<br>ug/L        | Acceptance<br>Limit | Conc<br>Qual     | LOD   | CRQL                  | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| CCB03                      | Thallium  | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 14:59            | LB135595      |
|                            | Vanadium  | 5.00                  | +/-5.00             | U                | 0.25  | 5.00                  | P | 04/29/2025       | 14:59            | LB135595      |
|                            | Zinc      | 5.00                  | +/-5.00             | U                | 1.50  | 5.00                  | P | 04/29/2025       | 14:59            | LB135595      |
| CCB04                      | Antimony  | 2.00                  | +/-2.00             | U                | 0.25  | 2.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Arsenic   | 1.00                  | +/-1.00             | U                | 0.25  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Barium    | 10.0                  | +/-10.0             | U                | 1.25  | 10.0                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Beryllium | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Cadmium   | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Chromium  | 2.00                  | +/-2.00             | U                | 0.75  | 2.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Copper    | 2.00                  | +/-2.00             | U                | 1.50  | 2.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Lead      | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Nickel    | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Selenium  | 5.00                  | +/-5.00             | U                | 4.50  | 5.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Silver    | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Thallium  | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Vanadium  | 5.00                  | +/-5.00             | U                | 0.25  | 5.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Zinc      | 5.00                  | +/-5.00             | U                | 1.50  | 5.00                  | P | 04/29/2025       | 15:48            | LB135595      |
|                            | Antimony  | 2.00                  | +/-2.00             | U                | 0.25  | 2.00                  | P | 04/29/2025       | 16:28            | LB135595      |
| CCB05                      | Arsenic   | 1.00                  | +/-1.00             | U                | 0.25  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Barium    | 10.0                  | +/-10.0             | U                | 1.25  | 10.0                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Beryllium | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Cadmium   | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Chromium  | 2.00                  | +/-2.00             | U                | 0.75  | 2.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Copper    | 2.00                  | +/-2.00             | U                | 1.50  | 2.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Lead      | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Nickel    | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Selenium  | 5.00                  | +/-5.00             | U                | 4.50  | 5.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Silver    | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Thallium  | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Vanadium  | 5.00                  | +/-5.00             | U                | 0.25  | 5.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Zinc      | 5.00                  | +/-5.00             | U                | 1.50  | 5.00                  | P | 04/29/2025       | 16:28            | LB135595      |
|                            | Antimony  | 0.12                  | +/-2.00             | J                | 0.25  | 2.00                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Arsenic   | 1.00                  | +/-1.00             | U                | 0.25  | 1.00                  | P | 04/29/2025       | 17:05            | LB135595      |
| CCB06                      | Barium    | 10.0                  | +/-10.0             | U                | 1.25  | 10.0                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Beryllium | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Cadmium   | 1.00                  | +/-1.00             | U                | 0.50  | 1.00                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Chromium  | 2.00                  | +/-2.00             | U                | 0.75  | 2.00                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Copper    | 2.00                  | +/-2.00             | U                | 1.50  | 2.00                  | P | 04/29/2025       | 17:05            | LB135595      |
|                            | Lead      | 1.00                  | +/-1.00             | U                | 0.75  | 1.00                  | P | 04/29/2025       | 17:05            | LB135595      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

| <b>Client:</b> <u>Nobis Group</u> |                              | <b>SDG No.:</b> <u>Q1883</u>  |                              |              |      |      |   |                  |                  |               |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|--------------|------|------|---|------------------|------------------|---------------|
| <b>Contract:</b> <u>NOBI03</u>    | <b>Lab Code:</b> <u>CHEM</u> | <b>Case No.:</b> <u>Q1883</u> | <b>SAS No.:</b> <u>Q1883</u> |              |      |      |   |                  |                  |               |
| Sample ID                         | Analyte                      | Result<br>ug/L                | Acceptance<br>Limit          | Conc<br>Qual | LOD  | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
| <b>CCB06</b>                      | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 17:05            | LB135595      |
|                                   | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 04/29/2025       | 17:05            | LB135595      |
|                                   | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 17:05            | LB135595      |
|                                   | Thallium                     | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 17:05            | LB135595      |
|                                   | Vanadium                     | 5.00                          | +/-5.00                      | U            | 0.25 | 5.00 | P | 04/29/2025       | 17:05            | LB135595      |
|                                   | Zinc                         | 5.00                          | +/-5.00                      | U            | 1.50 | 5.00 | P | 04/29/2025       | 17:05            | LB135595      |
| <b>CCB07</b>                      | Antimony                     | 2.00                          | +/-2.00                      | U            | 0.25 | 2.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Arsenic                      | 1.00                          | +/-1.00                      | U            | 0.25 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Barium                       | 10.0                          | +/-10.0                      | U            | 1.25 | 10.0 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Beryllium                    | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Cadmium                      | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Chromium                     | 2.00                          | +/-2.00                      | U            | 0.75 | 2.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Copper                       | 2.00                          | +/-2.00                      | U            | 1.50 | 2.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Lead                         | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Nickel                       | 1.00                          | +/-1.00                      | U            | 0.75 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Selenium                     | 5.00                          | +/-5.00                      | U            | 4.50 | 5.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Silver                       | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Thallium                     | 1.00                          | +/-1.00                      | U            | 0.50 | 1.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Vanadium                     | 5.00                          | +/-5.00                      | U            | 0.25 | 5.00 | P | 04/29/2025       | 17:50            | LB135595      |
|                                   | Zinc                         | 5.00                          | +/-5.00                      | U            | 1.50 | 5.00 | P | 04/29/2025       | 17:50            | LB135595      |