

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q1883
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1883 **SAS No.:** Q1883
Matrix: Water **Sample ID:** Q1883-16 **Client ID:** OU4-VSL-19-042325DUP
Percent Solids for Sample: NA **Duplicate ID** Q1883-16DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	10.0	UD	10.0	UD			P
Arsenic	ug/L	20	7.50	JD	4.00	JD	61		P
Barium	ug/L	20	119	D	119	D	0		P
Beryllium	ug/L	20	5.00	UD	5.00	UD			P
Cadmium	ug/L	20	5.00	UD	5.00	UD			P
Chromium	ug/L	20	16.4	D	16.5	D	1		P
Copper	ug/L	20	110	D	111	D	1		P
Lead	ug/L	20	6.85	D	3.35	JD	69		P
Nickel	ug/L	20	63.2	D	63.4	D	0		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	125	UD	125	UD			P
Silver	ug/L	20	5.00	UD	5.00	UD			P
Thallium	ug/L	20	5.00	UD	5.00	UD			P
Vanadium	ug/L	20	323	D	329	D	2		P
Zinc	ug/L	20	986	D	994	D	1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q1883
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1883 **SAS No.:** Q1883
Matrix: Water **Sample ID:** Q1883-16MS **Client ID:** OU4-VSL-19-042325MSD
Percent Solids for Sample: NA **Duplicate ID** Q1883-16MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	497	D	494	D	1		P
Arsenic	ug/L	20	4.25	JD	3.75	JD	13		P
Barium	ug/L	20	2360	D	2330	D	1		P
Beryllium	ug/L	20	506	D	507	D	0		P
Cadmium	ug/L	20	488	D	480	D	2		P
Chromium	ug/L	20	480	D	480	D	0		P
Copper	ug/L	20	5120	D	5130	D	0		P
Lead	ug/L	20	2330	D	2330	D	0		P
Nickel	ug/L	20	529	D	526	D	1		P
Mercury	ug/L	20	2.69		2.61		3		CV
Selenium	ug/L	20	461	D	473	D	3		P
Silver	ug/L	20	80.4	D	80.1	D	0		P
Thallium	ug/L	20	468	D	468	D	0		P
Vanadium	ug/L	20	795	D	792	D	0		P
Zinc	ug/L	20	5680	D	5660	D	0		P