

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1883
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1883 **SAS No.:** Q1883

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167780BS							
Antimony	ug/L	500	522		104	85 - 117	P
Arsenic	ug/L	500	519		104	84 - 116	P
Barium	ug/L	2500	2600		104	86 - 114	P
Beryllium	ug/L	500	540		108	83 - 121	P
Cadmium	ug/L	500	531		106	87 - 115	P
Chromium	ug/L	500	520		104	85 - 116	P
Copper	ug/L	5000	5110		102	85 - 118	P
Lead	ug/L	2500	2490		100	88 - 115	P
Nickel	ug/L	500	517		103	85 - 117	P
Selenium	ug/L	500	499		100	80 - 120	P
Silver	ug/L	500	526		105	85 - 116	P
Thallium	ug/L	500	500		100	82 - 116	P
Vanadium	ug/L	500	518		104	86 - 115	P
Zinc	ug/L	5000	5070		101	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1883
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1883 **SAS No.:** Q1883

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167789BS Mercury	ug/L	4.0	4.51		113	82 - 119	CV