

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 11:01 12:37

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.11	7.12	7.02	7.22	0.01
2,4-DCAA	6.94	6.94	6.84	7.04	0.00
Dalapon	2.45	2.45	2.35	2.55	0.00
DICHLORPROP	7.80	7.80	7.70	7.90	0.00
2,4-D	8.02	8.02	7.92	8.12	0.00
2,4,5-TP(Silvex)	8.86	8.86	8.76	8.96	0.00
2,4,5-T	9.14	9.14	9.04	9.24	0.00
2,4-DB	9.70	9.70	9.60	9.80	0.00
Dinoseb	10.86	10.87	10.77	10.97	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

Continuing Calib Date: 05/05/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 11:01 12:37

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.64	7.65	7.55	7.75	0.01
2,4-DCAA	7.46	7.47	7.37	7.57	0.01
Dalapon	2.52	2.53	2.43	2.63	0.01
DICHLORPROP	8.34	8.35	8.25	8.45	0.01
2,4-D	8.65	8.66	8.56	8.76	0.01
2,4,5-TP(Silvex)	9.53	9.54	9.44	9.64	0.01
2,4,5-T	9.93	9.95	9.85	10.05	0.02
2,4-DB	10.49	10.51	10.41	10.61	0.02
Dinoseb	10.87	10.88	10.78	10.98	0.02