

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029991.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.141	9.044	9.244	736.550	712.500	3.4
2,4,5-TP(Silvex)	8.858	8.762	8.962	730.550	712.500	2.5
2,4-D	8.016	7.919	8.119	715.650	705.000	1.5
2,4-DB	9.699	9.603	9.803	764.140	712.500	7.2
2,4-DCAA	6.939	6.842	7.042	758.550	750.000	1.1
Dalapon	2.449	2.349	2.549	706.400	682.500	3.5
DICAMBA	7.116	7.018	7.218	723.860	705.000	2.7
DICHLORPROP	7.796	7.699	7.899	704.420	705.000	-0.1
Dinoseb	10.861	10.766	10.966	714.800	705.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS029991.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.938	9.847	10.047	776.840	712.500	9.0
2,4,5-TP(Silvex)	9.534	9.442	9.642	786.260	712.500	10.4
2,4-D	8.655	8.561	8.761	748.240	705.000	6.1
2,4-DB	10.497	10.407	10.607	729.270	712.500	2.4
2,4-DCAA	7.463	7.367	7.567	801.050	750.000	6.8
Dalapon	2.523	2.425	2.625	711.230	682.500	4.2
DICAMBA	7.649	7.553	7.753	781.610	705.000	10.9
DICHLORPROP	8.342	8.248	8.448	757.570	705.000	7.5
Dinoseb	10.870	10.781	10.981	758.310	705.000	7.6