

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030002.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.140	9.044	9.244	722.200	712.500	1.4
2,4,5-TP(Silvex)	8.858	8.762	8.962	720.000	712.500	1.1
2,4-D	8.015	7.919	8.119	700.150	705.000	-0.7
2,4-DB	9.698	9.603	9.803	748.260	712.500	5.0
2,4-DCAA	6.939	6.842	7.042	735.570	750.000	-1.9
Dalapon	2.449	2.349	2.549	677.710	682.500	-0.7
DICAMBA	7.115	7.018	7.218	705.880	705.000	0.1
DICHLORPROP	7.796	7.699	7.899	687.850	705.000	-2.4
Dinoseb	10.860	10.766	10.966	690.110	705.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030002.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.938	9.847	10.047	770.350	712.500	8.1
2,4,5-TP(Silvex)	9.534	9.442	9.642	779.550	712.500	9.4
2,4-D	8.655	8.561	8.761	741.780	705.000	5.2
2,4-DB	10.496	10.407	10.607	739.380	712.500	3.8
2,4-DCAA	7.462	7.367	7.567	794.140	750.000	5.9
Dalapon	2.523	2.425	2.625	696.160	682.500	2.0
DICAMBA	7.648	7.553	7.753	780.920	705.000	10.8
DICHLORPROP	8.342	8.248	8.448	760.750	705.000	7.9
Dinoseb	10.870	10.781	10.981	739.550	705.000	4.9