

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030013.D Time Analyzed: 12:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.140	9.044	9.244	745.600	712.500	4.6
2,4,5-TP(Silvex)	8.858	8.762	8.962	742.010	712.500	4.1
2,4-D	8.015	7.919	8.119	720.060	705.000	2.1
2,4-DB	9.698	9.603	9.803	763.740	712.500	7.2
2,4-DCAA	6.939	6.842	7.042	760.450	750.000	1.4
Dalapon	2.450	2.349	2.549	694.070	682.500	1.7
DICAMBA	7.116	7.018	7.218	728.740	705.000	3.4
DICHLORPROP	7.796	7.699	7.899	708.330	705.000	0.5
Dinoseb	10.861	10.766	10.966	719.310	705.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030013.D Time Analyzed: 12:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.937	9.847	10.047	784.080	712.500	10.0
2,4,5-TP(Silvex)	9.533	9.442	9.642	796.690	712.500	11.8
2,4-D	8.654	8.561	8.761	757.120	705.000	7.4
2,4-DB	10.495	10.407	10.607	757.220	712.500	6.3
2,4-DCAA	7.462	7.367	7.567	820.590	750.000	9.4
Dalapon	2.523	2.425	2.625	705.560	682.500	3.4
DICAMBA	7.648	7.553	7.753	800.730	705.000	13.6
DICHLORPROP	8.341	8.248	8.448	768.250	705.000	9.0
Dinoseb	10.869	10.781	10.981	755.960	705.000	7.2