

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL04 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030016.D Time Analyzed: 13:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.139	9.044	9.244	754.230	712.500	5.9
2,4,5-TP(Silvex)	8.858	8.762	8.962	749.690	712.500	5.2
2,4-D	8.014	7.919	8.119	725.960	705.000	3.0
2,4-DB	9.698	9.603	9.803	768.520	712.500	7.9
2,4-DCAA	6.938	6.842	7.042	765.770	750.000	2.1
Dalapon	2.449	2.349	2.549	683.250	682.500	0.1
DICAMBA	7.115	7.018	7.218	731.570	705.000	3.8
DICHLORPROP	7.795	7.699	7.899	714.120	705.000	1.3
Dinoseb	10.860	10.766	10.966	724.710	705.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL04 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030016.D Time Analyzed: 13:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.937	9.847	10.047	784.510	712.500	10.1
2,4,5-TP(Silvex)	9.533	9.442	9.642	796.150	712.500	11.7
2,4-D	8.654	8.561	8.761	757.870	705.000	7.5
2,4-DB	10.495	10.407	10.607	769.370	712.500	8.0
2,4-DCAA	7.462	7.367	7.567	816.790	750.000	8.9
Dalapon	2.523	2.425	2.625	696.660	682.500	2.1
DICAMBA	7.648	7.553	7.753	796.150	705.000	12.9
DICHLORPROP	8.341	8.248	8.448	771.490	705.000	9.4
Dinoseb	10.869	10.781	10.981	758.940	705.000	7.7