

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL05 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030026.D Time Analyzed: 17:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.139	9.044	9.244	757.220	712.500	6.3
2,4,5-TP(Silvex)	8.857	8.762	8.962	751.430	712.500	5.5
2,4-D	8.015	7.919	8.119	727.280	705.000	3.2
2,4-DB	9.698	9.603	9.803	769.080	712.500	7.9
2,4-DCAA	6.938	6.842	7.042	767.300	750.000	2.3
Dalapon	2.448	2.349	2.549	682.430	682.500	0.0
DICAMBA	7.114	7.018	7.218	731.810	705.000	3.8
DICHLORPROP	7.795	7.699	7.899	717.350	705.000	1.8
Dinoseb	10.860	10.766	10.966	739.100	705.000	4.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL05 Date Analyzed: 05/01/2025

Lab Sample No.: HSTDCCC750 Data File : PS030026.D Time Analyzed: 17:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.937	9.847	10.047	790.510	712.500	10.9
2,4,5-TP(Silvex)	9.533	9.442	9.642	801.340	712.500	12.5
2,4-D	8.653	8.561	8.761	765.270	705.000	8.5
2,4-DB	10.495	10.407	10.607	767.700	712.500	7.7
2,4-DCAA	7.462	7.367	7.567	808.570	750.000	7.8
Dalapon	2.523	2.425	2.625	682.630	682.500	0.0
DICAMBA	7.648	7.553	7.753	797.610	705.000	13.1
DICHLORPROP	8.341	8.248	8.448	798.240	705.000	13.2
Dinoseb	10.869	10.781	10.981	761.200	705.000	8.0