

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL06 Date Analyzed: 05/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030029.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.140	9.044	9.244	757.080	712.500	6.3
2,4,5-TP(Silvex)	8.857	8.762	8.962	758.140	712.500	6.4
2,4-D	8.015	7.919	8.119	776.810	705.000	10.2
2,4-DB	9.698	9.603	9.803	770.300	712.500	8.1
2,4-DCAA	6.938	6.842	7.042	772.860	750.000	3.0
Dalapon	2.446	2.349	2.549	769.750	682.500	12.8
DICAMBA	7.114	7.018	7.218	745.680	705.000	5.8
DICHLORPROP	7.795	7.699	7.899	758.480	705.000	7.6
Dinoseb	10.861	10.766	10.966	741.300	705.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL06 Date Analyzed: 05/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030029.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.933	9.847	10.047	770.140	712.500	8.1
2,4,5-TP(Silvex)	9.528	9.442	9.642	784.120	712.500	10.1
2,4-D	8.649	8.561	8.761	745.960	705.000	5.8
2,4-DB	10.491	10.407	10.607	757.060	712.500	6.3
2,4-DCAA	7.456	7.367	7.567	788.400	750.000	5.1
Dalapon	2.517	2.425	2.625	730.340	682.500	7.0
DICAMBA	7.642	7.553	7.753	789.280	705.000	12.0
DICHLORPROP	8.336	8.248	8.448	749.950	705.000	6.4
Dinoseb	10.865	10.781	10.981	753.620	705.000	6.9