

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL07 Date Analyzed: 05/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030041.D Time Analyzed: 16:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.139	9.044	9.244	707.160	712.500	-0.7
2,4,5-TP(Silvex)	8.857	8.762	8.962	704.000	712.500	-1.2
2,4-D	8.014	7.919	8.119	687.970	705.000	-2.4
2,4-DB	9.699	9.603	9.803	717.320	712.500	0.7
2,4-DCAA	6.936	6.842	7.042	721.190	750.000	-3.8
Dalapon	2.445	2.349	2.549	720.930	682.500	5.6
DICAMBA	7.114	7.018	7.218	688.440	705.000	-2.3
DICHLORPROP	7.794	7.699	7.899	674.770	705.000	-4.3
Dinoseb	10.861	10.766	10.966	699.240	705.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL07 Date Analyzed: 05/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030041.D Time Analyzed: 16:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.932	9.847	10.047	719.530	712.500	1.0
2,4,5-TP(Silvex)	9.527	9.442	9.642	729.410	712.500	2.4
2,4-D	8.648	8.561	8.761	696.650	705.000	-1.2
2,4-DB	10.490	10.407	10.607	693.040	712.500	-2.7
2,4-DCAA	7.455	7.367	7.567	737.500	750.000	-1.7
Dalapon	2.516	2.425	2.625	674.830	682.500	-1.1
DICAMBA	7.642	7.553	7.753	730.790	705.000	3.7
DICHLORPROP	8.335	8.248	8.448	708.150	705.000	0.4
Dinoseb	10.864	10.781	10.981	696.840	705.000	-1.2