

Surrogate Summary

SDG No.: Q1883

Client: Nobis Group

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029918.D	PIBLK-PS029918.D	2,4-DCAA	1	500	433	87		32	138
		2,4-DCAA	2	500	433	87		32	138
I.BLK-PS029990.D	PIBLK-PS029990.D	2,4-DCAA	1	500	525	105		32	138
		2,4-DCAA	2	500	558	112		32	138
Q1906-05MS	WC-5MS	2,4-DCAA	1	500	956	191	*	27	122
		2,4-DCAA	2	500	1520	303	*	27	122
Q1906-05MSD	WC-5MSD	2,4-DCAA	1	500	969	194	*	27	122
		2,4-DCAA	2	500	1540	308	*	27	122
I.BLK-PS030001.D	PIBLK-PS030001.D	2,4-DCAA	1	500	518	104		32	138
		2,4-DCAA	2	500	565	113		32	138
PB167796BL	PB167796BL	2,4-DCAA	1	500	433	87		27	122
		2,4-DCAA	2	500	413	83		27	122
I.BLK-PS030012.D	PIBLK-PS030012.D	2,4-DCAA	1	500	551	110		32	138
		2,4-DCAA	2	500	588	118		32	138
I.BLK-PS030015.D	PIBLK-PS030015.D	2,4-DCAA	1	500	534	107		32	138
		2,4-DCAA	2	500	567	113		32	138
Q1883-01	OU4-PCS-TC-27-042325	2,4-DCAA	1	500	412	82		27	122
		2,4-DCAA	2	500	400	80		27	122
Q1883-03	OU4-PCS-TC-28-042325	2,4-DCAA	1	500	338	68		27	122
		2,4-DCAA	2	500	337	67		27	122
Q1883-05	OU4-PCS-TC-29-042325	2,4-DCAA	1	500	138	28		27	122
		2,4-DCAA	2	500	135	27		27	122
Q1883-07	OU4-PCS-TC-30-042325	2,4-DCAA	1	500	371	74		27	122
		2,4-DCAA	2	500	350	70		27	122
Q1883-09	OU4-PCS-TC-31-042325	2,4-DCAA	1	500	390	78		27	122
		2,4-DCAA	2	500	380	76		27	122
Q1883-11	OU4-PCS-TC-32-042325	2,4-DCAA	1	500	327	65		27	122
		2,4-DCAA	2	500	318	64		27	122
Q1883-13	OU4-VSL-18-042325	2,4-DCAA	1	500	365	73		27	122
		2,4-DCAA	2	500	345	69		27	122
Q1883-15	OU4-VSL-19-042325	2,4-DCAA	1	500	401	80		27	122
		2,4-DCAA	2	500	409	82		27	122
I.BLK-PS030025.D	PIBLK-PS030025.D	2,4-DCAA	1	500	562	112		32	138
		2,4-DCAA	2	500	598	120		32	138
I.BLK-PS030028.D	PIBLK-PS030028.D	2,4-DCAA	1	500	530	106		32	138
		2,4-DCAA	2	500	553	111		32	138
PB167796BS	PB167796BS	2,4-DCAA	1	500	530	106		27	122
		2,4-DCAA	2	500	540	108		27	122
I.BLK-PS030040.D	PIBLK-PS030040.D	2,4-DCAA	1	500	529	106		32	138
		2,4-DCAA	2	500	552	110		32	138