

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.679	4.879	567.510	500.000	13.5
Aroclor-1016-2	4.795	4.698	4.898	567.550	500.000	13.5
Aroclor-1016-3	4.851	4.755	4.955	552.160	500.000	10.4
Aroclor-1016-4	4.972	4.875	5.075	566.160	500.000	13.2
Aroclor-1016-5	5.229	5.132	5.332	568.900	500.000	13.8
Aroclor-1260-1	6.268	6.172	6.372	552.510	500.000	10.5
Aroclor-1260-2	6.457	6.360	6.560	540.900	500.000	8.2
Aroclor-1260-3	6.825	6.729	6.929	541.450	500.000	8.3
Aroclor-1260-4	7.084	6.988	7.188	532.720	500.000	6.5
Aroclor-1260-5	7.327	7.230	7.430	535.700	500.000	7.1
Decachlorobiphenyl	8.728	8.632	8.832	49.210	50.000	-1.6
Tetrachloro-m-xylene	3.686	3.588	3.788	57.430	50.000	14.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	533.100	500.000	6.6
Aroclor-1016-2	4.781	4.684	4.884	531.290	500.000	6.3
Aroclor-1016-3	4.956	4.860	5.060	513.330	500.000	2.7
Aroclor-1016-4	4.998	4.902	5.102	496.070	500.000	-0.8
Aroclor-1016-5	5.211	5.115	5.315	510.550	500.000	2.1
Aroclor-1260-1	6.242	6.146	6.346	502.200	500.000	0.4
Aroclor-1260-2	6.429	6.334	6.534	495.650	500.000	-0.9
Aroclor-1260-3	6.582	6.487	6.687	487.880	500.000	-2.4
Aroclor-1260-4	7.053	6.957	7.157	485.750	500.000	-2.9
Aroclor-1260-5	7.294	7.197	7.397	488.550	500.000	-2.3
Decachlorobiphenyl	8.679	8.584	8.784	46.590	50.000	-6.8
Tetrachloro-m-xylene	3.683	3.586	3.786	54.630	50.000	9.3