

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110876.D Time Analyzed: 22:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	557.860	500.000	11.6
Aroclor-1016-2	4.795	4.698	4.898	559.810	500.000	12.0
Aroclor-1016-3	4.851	4.755	4.955	538.720	500.000	7.7
Aroclor-1016-4	4.971	4.875	5.075	545.640	500.000	9.1
Aroclor-1016-5	5.229	5.132	5.332	551.180	500.000	10.2
Aroclor-1260-1	6.269	6.172	6.372	518.480	500.000	3.7
Aroclor-1260-2	6.457	6.360	6.560	510.870	500.000	2.2
Aroclor-1260-3	6.825	6.729	6.929	503.660	500.000	0.7
Aroclor-1260-4	7.085	6.988	7.188	490.870	500.000	-1.8
Aroclor-1260-5	7.327	7.230	7.430	489.900	500.000	-2.0
Decachlorobiphenyl	8.728	8.632	8.832	45.530	50.000	-8.9
Tetrachloro-m-xylene	3.686	3.588	3.788	57.090	50.000	14.2

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110876.D Time Analyzed: 22:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	524.970	500.000	5.0
Aroclor-1016-2	4.781	4.684	4.884	531.970	500.000	6.4
Aroclor-1016-3	4.957	4.860	5.060	500.250	500.000	0.1
Aroclor-1016-4	4.998	4.902	5.102	481.510	500.000	-3.7
Aroclor-1016-5	5.211	5.115	5.315	505.850	500.000	1.2
Aroclor-1260-1	6.243	6.146	6.346	482.790	500.000	-3.4
Aroclor-1260-2	6.431	6.334	6.534	475.660	500.000	-4.9
Aroclor-1260-3	6.583	6.487	6.687	468.070	500.000	-6.4
Aroclor-1260-4	7.054	6.957	7.157	461.140	500.000	-7.8
Aroclor-1260-5	7.295	7.197	7.397	461.730	500.000	-7.7
Decachlorobiphenyl	8.680	8.584	8.784	44.710	50.000	-10.6
Tetrachloro-m-xylene	3.683	3.586	3.786	54.470	50.000	8.9