

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071599.D Time Analyzed: 10:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.663	5.571	5.771	514.370	500.000	2.9
Aroclor-1016-2	5.685	5.593	5.793	504.710	500.000	0.9
Aroclor-1016-3	5.747	5.654	5.854	510.290	500.000	2.1
Aroclor-1016-4	5.844	5.752	5.952	520.150	500.000	4.0
Aroclor-1016-5	6.137	6.045	6.245	506.180	500.000	1.2
Aroclor-1260-1	7.256	7.165	7.365	529.010	500.000	5.8
Aroclor-1260-2	7.509	7.418	7.618	515.870	500.000	3.2
Aroclor-1260-3	7.868	7.777	7.977	531.680	500.000	6.3
Aroclor-1260-4	8.092	8.002	8.202	513.780	500.000	2.8
Aroclor-1260-5	8.412	8.322	8.522	529.980	500.000	6.0
Decachlorobiphenyl	10.225	10.137	10.337	54.140	50.000	8.3
Tetrachloro-m-xylene	4.510	4.417	4.617	50.740	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071599.D Time Analyzed: 10:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.892	4.798	4.998	525.310	500.000	5.1
Aroclor-1016-2	4.911	4.816	5.016	530.190	500.000	6.0
Aroclor-1016-3	5.088	4.993	5.193	543.770	500.000	8.8
Aroclor-1016-4	5.130	5.036	5.236	528.260	500.000	5.7
Aroclor-1016-5	5.344	5.250	5.450	571.750	500.000	14.4
Aroclor-1260-1	6.378	6.285	6.485	556.170	500.000	11.2
Aroclor-1260-2	6.566	6.473	6.673	550.380	500.000	10.1
Aroclor-1260-3	6.720	6.627	6.827	543.580	500.000	8.7
Aroclor-1260-4	7.190	7.098	7.298	548.770	500.000	9.8
Aroclor-1260-5	7.432	7.339	7.539	538.670	500.000	7.7
Decachlorobiphenyl	8.840	8.749	8.949	53.370	50.000	6.7
Tetrachloro-m-xylene	3.808	3.712	3.912	56.790	50.000	13.6