

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071629.D Time Analyzed: 21:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	532.730	500.000	6.5
Aroclor-1016-2	5.687	5.593	5.793	523.890	500.000	4.8
Aroclor-1016-3	5.749	5.654	5.854	524.180	500.000	4.8
Aroclor-1016-4	5.846	5.752	5.952	526.580	500.000	5.3
Aroclor-1016-5	6.139	6.045	6.245	533.550	500.000	6.7
Aroclor-1260-1	7.257	7.165	7.365	548.960	500.000	9.8
Aroclor-1260-2	7.511	7.418	7.618	534.240	500.000	6.8
Aroclor-1260-3	7.870	7.777	7.977	546.610	500.000	9.3
Aroclor-1260-4	8.094	8.002	8.202	533.340	500.000	6.7
Aroclor-1260-5	8.413	8.322	8.522	548.400	500.000	9.7
Decachlorobiphenyl	10.228	10.137	10.337	55.780	50.000	11.6
Tetrachloro-m-xylene	4.511	4.417	4.617	52.610	50.000	5.2

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071629.D Time Analyzed: 21:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	544.770	500.000	9.0
Aroclor-1016-2	4.912	4.816	5.016	551.650	500.000	10.3
Aroclor-1016-3	5.089	4.993	5.193	561.570	500.000	12.3
Aroclor-1016-4	5.130	5.036	5.236	543.490	500.000	8.7
Aroclor-1016-5	5.345	5.250	5.450	582.730	500.000	16.5
Aroclor-1260-1	6.379	6.285	6.485	545.780	500.000	9.2
Aroclor-1260-2	6.568	6.473	6.673	539.920	500.000	8.0
Aroclor-1260-3	6.721	6.627	6.827	531.780	500.000	6.4
Aroclor-1260-4	7.192	7.098	7.298	525.410	500.000	5.1
Aroclor-1260-5	7.434	7.339	7.539	525.950	500.000	5.2
Decachlorobiphenyl	8.841	8.749	8.949	54.580	50.000	9.2
Tetrachloro-m-xylene	3.809	3.712	3.912	55.200	50.000	10.4