

Surrogate Summary

SDG No.: Q1883

Client: Nobis Group

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095326.D	PIBLK-PL095326.D	Decachlorobiphenyl	1	20	18.9	94		30	135
		Tetrachloro-m-xylene	1	20	17.3	87		44	124
		Decachlorobiphenyl	2	20	18.2	91		30	135
		Tetrachloro-m-xylene	2	20	17.1	86		44	124
I.BLK-PL095438.D	PIBLK-PL095438.D	Decachlorobiphenyl	1	20	22.9	114		30	135
		Tetrachloro-m-xylene	1	20	18.6	93		44	124
		Decachlorobiphenyl	2	20	21.7	109		30	135
		Tetrachloro-m-xylene	2	20	17.7	89		44	124
PB167766BS	PB167766BS	Decachlorobiphenyl	1	20	23.6	118		55	130
		Tetrachloro-m-xylene	1	20	20.7	104		42	129
		Decachlorobiphenyl	2	20	23.8	119		55	130
		Tetrachloro-m-xylene	2	20	20.8	104		42	129
Q1889-01MS	COMP-1MS	Decachlorobiphenyl	1	20	17.6	88		55	130
		Tetrachloro-m-xylene	1	20	18.1	90		42	129
		Decachlorobiphenyl	2	20	17.8	89		55	130
		Tetrachloro-m-xylene	2	20	18.4	92		42	129
Q1889-01MSD	COMP-1MSD	Decachlorobiphenyl	1	20	17.4	87		55	130
		Tetrachloro-m-xylene	1	20	17.9	90		42	129
		Decachlorobiphenyl	2	20	17.1	86		55	130
		Tetrachloro-m-xylene	2	20	18.0	90		42	129
I.BLK-PL095450.D	PIBLK-PL095450.D	Decachlorobiphenyl	1	20	21.2	106		30	135
		Tetrachloro-m-xylene	1	20	18.5	93		44	124
		Decachlorobiphenyl	2	20	21.8	109		30	135
		Tetrachloro-m-xylene	2	20	19.0	95		44	124
I.BLK-PL095473.D	PIBLK-PL095473.D	Decachlorobiphenyl	1	20	22.2	111		30	135
		Tetrachloro-m-xylene	1	20	18.9	95		44	124
		Decachlorobiphenyl	2	20	21.0	105		30	135
		Tetrachloro-m-xylene	2	20	18.1	91		44	124
PB167766BL	PB167766BL	Decachlorobiphenyl	1	20	21.5	107		55	130
		Tetrachloro-m-xylene	1	20	19.4	97		42	129
		Decachlorobiphenyl	2	20	20.9	105		55	130
		Tetrachloro-m-xylene	2	20	18.4	92		42	129
Q1883-01	OU4-PCS-TC-27-042325	Decachlorobiphenyl	1	20	20.1	101		55	130
		Tetrachloro-m-xylene	1	20	18.9	95		42	129
		Decachlorobiphenyl	2	20	18.9	95		55	130
		Tetrachloro-m-xylene	2	20	21.1	106		42	129
I.BLK-PL095483.D	PIBLK-PL095483.D	Decachlorobiphenyl	1	20	22.9	115		30	135
		Tetrachloro-m-xylene	1	20	19.1	96		44	124
		Decachlorobiphenyl	2	20	21.8	109		30	135
		Tetrachloro-m-xylene	2	20	21.8	109		44	124
Q1883-03	OU4-PCS-TC-28-042325	Decachlorobiphenyl	1	20	20.4	102		55	130

Surrogate Summary

SDG No.: **Q1883**

Client: **Nobis Group**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1883-03	OU4-PCS-TC-28-042325	Tetrachloro-m-xylene	1	20	19.1	95		42	129
		Decachlorobiphenyl	2	20	21.7	108		55	130
		Tetrachloro-m-xylene	2	20	23.0	115		42	129
Q1883-05	OU4-PCS-TC-29-042325	Decachlorobiphenyl	1	20	14.4	72		55	130
		Tetrachloro-m-xylene	1	20	14.4	72		42	129
		Decachlorobiphenyl	2	20	14.4	72		55	130
		Tetrachloro-m-xylene	2	20	18.1	91		42	129
Q1883-07	OU4-PCS-TC-30-042325	Decachlorobiphenyl	1	20	20.3	102		55	130
		Tetrachloro-m-xylene	1	20	18.1	91		42	129
		Decachlorobiphenyl	2	20	19.5	97		55	130
		Tetrachloro-m-xylene	2	20	21.8	109		42	129
Q1883-09	OU4-PCS-TC-31-042325	Decachlorobiphenyl	1	20	20.9	104		55	130
		Tetrachloro-m-xylene	1	20	19.1	96		42	129
		Decachlorobiphenyl	2	20	20.1	100		55	130
		Tetrachloro-m-xylene	2	20	21.5	107		42	129
Q1883-11	OU4-PCS-TC-32-042325	Decachlorobiphenyl	1	20	20.2	101		55	130
		Tetrachloro-m-xylene	1	20	17.7	88		42	129
		Decachlorobiphenyl	2	20	20.2	101		55	130
		Tetrachloro-m-xylene	2	20	21.0	105		42	129
Q1883-13	OU4-VSL-18-042325	Decachlorobiphenyl	1	20	20.8	104		55	130
		Tetrachloro-m-xylene	1	20	19.8	99		42	129
		Decachlorobiphenyl	2	20	21.0	105		55	130
		Tetrachloro-m-xylene	2	20	24.3	122		42	129
Q1883-15	OU4-VSL-19-042325	Decachlorobiphenyl	1	20	20.3	102		55	130
		Tetrachloro-m-xylene	1	20	19.0	95		42	129
		Decachlorobiphenyl	2	20	20.5	103		55	130
		Tetrachloro-m-xylene	2	20	23.6	118		42	129
I.BLK-PL095494.D	PIBLK-PL095494.D	Decachlorobiphenyl	1	20	21.6	108		30	135
		Tetrachloro-m-xylene	1	20	18.7	93		44	124
		Decachlorobiphenyl	2	20	21.8	109		30	135
		Tetrachloro-m-xylene	2	20	23.5	118		44	124