

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM042825

Review By	Rahul	Review On	4/29/2025 8:56:26 AM
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:30 AM
SubDirectory	BM042825	HP Acquire Method	BNA_M
		HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050023.D	28 Apr 2025 11:46		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050024.D	28 Apr 2025 12:30		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050025.D	28 Apr 2025 13:09	Compound#32-41-54-56-65-70-7 7 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM050026.D	28 Apr 2025 13:48		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050027.D	28 Apr 2025 14:27		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050028.D	28 Apr 2025 15:06	Compound#54 & 56 Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BM050029.D	28 Apr 2025 15:45		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BM050030.D	28 Apr 2025 16:24		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050031.D	28 Apr 2025 17:04		RC/JU	Ok
10	SSTDICV040	ICVBM042825	BM050032.D	28 Apr 2025 17:43		RC/JU	Ok,M
11	PB167739BL	PB167739BL	BM050033.D	28 Apr 2025 19:40	Analyzed for contamination check	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM043025

Review By	Rahul	Review On	5/1/2025 11:29:43 AM
Supervise By	Jagrut	Supervise On	5/1/2025 1:18:51 PM
SubDirectory	BM043025	HP Acquire Method	BNA_M
		HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050041.D	30 Apr 2025 10:14		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050042.D	30 Apr 2025 10:53		RC/JU	Ok
3	PB167767BL	PB167767BL	BM050043.D	30 Apr 2025 11:32	Not Used	RC/JU	Not Ok
4	Q1883-01	OU4-PCS-TC-27-0423	BM050044.D	30 Apr 2025 14:07		RC/JU	Ok
5	Q1883-03	OU4-PCS-TC-28-0423	BM050045.D	30 Apr 2025 14:46		RC/JU	Ok
6	Q1883-05	OU4-PCS-TC-29-0423	BM050046.D	30 Apr 2025 15:25	Surrogate Fail - RE EXTRACTION.	RC/JU	ReRun
7	Q1883-07	OU4-PCS-TC-30-0423	BM050047.D	30 Apr 2025 16:04		RC/JU	Ok
8	Q1883-09	OU4-PCS-TC-31-0423	BM050048.D	30 Apr 2025 16:44		RC/JU	Ok
9	Q1889-02	COMP-2	BM050049.D	30 Apr 2025 17:23		RC/JU	Ok
10	Q1889-02MS	COMP-2MS	BM050050.D	30 Apr 2025 18:02		RC/JU	Ok
11	Q1889-02MSD	COMP-2MSD	BM050051.D	30 Apr 2025 18:41		RC/JU	Ok
12	SSTDCCC040	SSTDCCC040EC	BM050052.D	30 Apr 2025 19:20		RC/JU	Ok
13	DFTPP	DFTPP	BM050053.D	30 Apr 2025 20:38		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BM050054.D	30 Apr 2025 21:18		RC/JU	Ok
15	PB167767BL	PB167767BL	BM050055.D	30 Apr 2025 21:57		RC/JU	Ok
16	PB167767BS	PB167767BS	BM050056.D	30 Apr 2025 22:36		RC/JU	Ok
17	Q1883-11	OU4-PCS-TC-32-0423	BM050057.D	30 Apr 2025 23:15		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM043025

Review By	Rahul	Review On	5/1/2025 11:29:43 AM		
Supervise By	Jagrut	Supervise On	5/1/2025 1:18:51 PM		
SubDirectory	BM043025	HP Acquire Method	BNA_M	HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q1883-13	OU4-VSL-18-042325	BM050058.D	30 Apr 2025 23:54		RC/JU	Ok
19	Q1883-15	OU4-VSL-19-042325	BM050059.D	01 May 2025 00:33		RC/JU	Ok
20	SSTDCCC040	SSTDCCC040EC	BM050060.D	01 May 2025 01:12		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM
SubDirectory	BP041425	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024274.D	14 Apr 2025 10:25		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024276.D	14 Apr 2025 11:47	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024277.D	14 Apr 2025 12:27		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024278.D	14 Apr 2025 13:08		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	This calibration is fail for Com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024280.D	14 Apr 2025 15:10	This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024281.D	14 Apr 2025 16:32	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024282.D	14 Apr 2025 17:13	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP041425	BP024283.D	14 Apr 2025 18:35		RC/JU	Ok,M
11	PB167488TB	PB167488TB	BP024284.D	14 Apr 2025 19:16		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP050525

Review By	Rahul	Review On	5/6/2025 9:58:11 AM
Supervise By	Jagrut	Supervise On	5/6/2025 10:50:58 AM
SubDirectory	BP050525	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024521.D	05 May 2025 11:57		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024522.D	05 May 2025 12:38		RC/JU	Ok
3	PB167836BL	PB167836BL	BP024523.D	05 May 2025 13:18		RC/JU	Ok
4	PB167836BS	PB167836BS	BP024524.D	05 May 2025 13:59		RC/JU	Ok
5	Q1883-05RX	OU4-PCS-TC-29-0423	BP024525.D	05 May 2025 14:43	Surrogate Fail	RC/JU	Confirms
6	Q1928-01MS	MH-QMS	BP024526.D	05 May 2025 15:24		RC/JU	Ok
7	Q1928-01MSD	MH-QMSD	BP024527.D	05 May 2025 16:05		RC/JU	Ok,M
8	Q1935-01	MH-NN	BP024528.D	05 May 2025 16:45		RC/JU	Ok
9	Q1901-10RE	B-167-SB02RE	BP024529.D	05 May 2025 17:26	Internal Standard and surrogate Fail	RC/JU	Confirms
10	Q1907-02RE	CO-8R-WCRE	BP024530.D	05 May 2025 18:07	Internal Standard and surrogate Fail	RC/JU	Confirms
11	Q1932-06RE	COMP-6RE	BP024531.D	05 May 2025 18:47	Surrogate Fail	RC/JU	Confirms
12	Q1929-07	WC-A1-03-C	BP024532.D	05 May 2025 19:28		RC/JU	Ok
13	Q1929-11	WC-A1-04-C	BP024533.D	05 May 2025 20:09		RC/JU	Ok
14	Q1935-04	MH-NN	BP024534.D	05 May 2025 20:49		RC/JU	Ok
15	Q1935-08	MH-MM	BP024535.D	05 May 2025 21:30		RC/JU	Ok
16	Q1928-04	MH-Q	BP024536.D	05 May 2025 22:11		RC/JU	Ok
17	Q1946-01	BB-CONCRETE	BP024537.D	05 May 2025 22:52		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP050525

Review By	Rahul	Review On	5/6/2025 9:58:11 AM		
Supervise By	Jagrut	Supervise On	5/6/2025 10:50:58 AM		
SubDirectory	BP050525	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	SSTDCCC040	SSTDCCC040EC	BP024538.D	05 May 2025 23:32		RC/JU	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration