

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/30/2025

Lab File ID: BM050054.D Time Analyzed: 21:18

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	250618	7.763	921090	10.56	604340	14.40
UPPER LIMIT	501236	8.263	1842180	11.057	1208680	14.904
LOWER LIMIT	125309	7.263	460545	10.057	302170	13.904
EPA SAMPLE NO.						
01 PB167767BL	259374	7.76	884512	10.56	572177	14.41
02 PB167767BS	260866	7.76	920346	10.56	574035	14.40
03 OU4-PCS-TC-32-042325	255652	7.76	890304	10.56	608919	14.41
04 OU4-VSL-18-042325	244720	7.76	862862	10.56	592525	14.41
05 OU4-VSL-19-042325	245309	7.76	863965	10.56	585954	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.