

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/07/2025

Lab File ID: BP024558.D Time Analyzed: 10:08

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	212497	7.699	880474	10.47	543552	14.33
UPPER LIMIT	424994	8.199	1760950	10.969	1087100	14.834
LOWER LIMIT	106249	7.199	440237	9.969	271776	13.834
EPA SAMPLE NO.						
01 MH-QMS	148768	7.70	581624	10.48	368354	14.34
02 MH-QMSD	157526	7.70	621901	10.48	393833	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.