

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: NOBI03

Lab Code: CHEM

SAS No.: Q1883

SDG NO.: Q1883

Lab File ID: BM050053.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_M

DFTPP Injection Time: 20:38

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.6
68	Less than 2.0% of mass 69	0.4 (1.5) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	34.3
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	26.7
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	13.3
442	Greater than 50% of mass 198	84.6
443	15.0 - 24.0% of mass 442	16.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050054.D	04/30/2025	21:18
PB167767BL	PB167767BL	BM050055.D	04/30/2025	21:57
PB167767BS	PB167767BS	BM050056.D	04/30/2025	22:36
OU4-PCS-TC-32-042325	Q1883-11	BM050057.D	04/30/2025	23:15
OU4-VSL-18-042325	Q1883-13	BM050058.D	04/30/2025	23:54
OU4-VSL-19-042325	Q1883-15	BM050059.D	05/01/2025	00:33
SSTDCCC040EC	SSTDCCC040	BM050060.D	05/01/2025	01:12