

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG No.: Q1883
Instrument ID: MSVOA_Y Calibration Date(s): 04/22/2025 04/22/2025
Heated Purge: (Y/N) Y Calibration Time(s): 13:39 16:15
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D	
COMPOUND		RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
Dichlorodifluoromethane		0.469	0.477	0.405	0.393	0.408	0.405	0.426	8.7				
Chloromethane		0.704	0.670	0.636	0.602	0.529	0.498	0.607	13.2				
Vinyl Chloride		0.829	0.830	0.758	0.730	0.680	0.647	0.745	10.1				
Bromomethane		0.782	0.697	0.674	0.596	0.554	0.550	0.642	14.3				
Chloroethane		0.589	0.550	0.524	0.494	0.460	0.429	0.508	11.6				
Tetrahydrofuran		0.067	0.065	0.072	0.073	0.066	0.067	0.068	4.9				
Trichlorofluoromethane		1.048	1.075	0.987	0.960	0.921	0.907	0.983	6.9				
1,1,2-Trichlorotrifluoroethane		0.604	0.591	0.530	0.489	0.497	0.488	0.533	9.8				
1,1-Dichloroethene		0.525	0.532	0.483	0.471	0.487	0.484	0.497	5				
Acrylonitrile		0.103	0.090	0.099	0.093	0.088	0.088	0.094	6.6				
Acetone		0.092	0.084	0.066	0.063	0.058	0.057	0.070	20.7				
Carbon Disulfide		1.736	1.737	1.569	1.493	1.516	1.459	1.585	7.7				
Methyl tert-butyl Ether		1.098	1.067	1.102	1.133	1.141	1.134	1.113	2.6				
Methylene Chloride		0.759	0.625	0.574	0.533	0.518	0.494	0.584	16.7				
trans-1,2-Dichloroethene		0.619	0.578	0.533	0.532	0.546	0.532	0.557	6.3				
1,1-Dichloroethane		0.987	0.943	0.893	0.853	0.855	0.826	0.893	6.9				
2-Butanone		0.111	0.107	0.108	0.106	0.099	0.100	0.105	4.5				
Carbon Tetrachloride		0.540	0.563	0.512	0.507	0.535	0.523	0.530	3.9				
2,2-Dichloropropane		0.841	0.830	0.763	0.743	0.770	0.748	0.782	5.4				
cis-1,2-Dichloroethene		0.626	0.638	0.604	0.612	0.633	0.621	0.622	2.1				
Chloroform		1.084	1.046	0.974	0.956	0.958	0.925	0.990	6.2				
1,1,1-Trichloroethane		0.972	0.944	0.896	0.853	0.865	0.846	0.896	5.8				
1,1-Dichloropropene		0.455	0.470	0.424	0.425	0.450	0.437	0.444	4.1				
Benzene		1.423	1.439	1.351	1.337	1.415	1.358	1.387	3.1				
1,2-Dichloroethane		0.347	0.367	0.350	0.335	0.336	0.325	0.343	4.3				
Trichloroethene		0.402	0.405	0.369	0.367	0.387	0.375	0.384	4.3				
1,2-Dichloropropane		0.320	0.317	0.302	0.300	0.308	0.294	0.307	3.4				
Dibromomethane		0.202	0.193	0.192	0.187	0.191	0.187	0.192	3				
Bromodichloromethane		0.479	0.504	0.464	0.470	0.490	0.471	0.480	3.1				
4-Methyl-2-Pentanone		0.144	0.149	0.167	0.172	0.166	0.165	0.160	6.9				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG No.: Q1883
 Instrument ID: MSVOA_Y Calibration Date(s): 04/22/2025 04/22/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 13:39 16:15
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D	
COMPOUND		RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
Toluene		0.838	0.915	0.875	0.894	0.952	0.913	0.898	4.4				
t-1,3-Dichloropropene		0.383	0.408	0.399	0.417	0.437	0.425	0.411	4.7				
cis-1,3-Dichloropropene		0.466	0.488	0.483	0.484	0.514	0.500	0.489	3.3				
1,1,2-Trichloroethane		0.250	0.261	0.261	0.256	0.258	0.252	0.256	1.8				
1,3-Dichloropropane		0.413	0.420	0.405	0.409	0.417	0.404	0.411	1.6				
2-Hexanone		0.093	0.096	0.109	0.115	0.111	0.111	0.106	8.5				
Dibromochloromethane		0.345	0.355	0.354	0.352	0.367	0.357	0.355	2				
1,2-Dibromoethane		0.238	0.243	0.243	0.241	0.247	0.240	0.242	1.3				
Tetrachloroethene		0.500	0.516	0.462	0.455	0.466	0.430	0.472	6.6				
Chlorobenzene		1.188	1.152	1.055	1.083	1.139	1.092	1.118	4.5				
1,1,1,2-Tetrachloroethane		0.401	0.409	0.376	0.386	0.401	0.388	0.394	3.1				
Ethyl Benzene		1.693	1.840	1.718	1.835	1.990	1.898	1.829	6.1				
m/p-Xylenes		0.664	0.720	0.699	0.734	0.797	0.754	0.728	6.3				
o-Xylene		0.599	0.639	0.635	0.689	0.747	0.716	0.671	8.3				
Styrene		0.950	1.080	1.078	1.169	1.272	1.209	1.126	10.2				
Bromoform		0.234	0.221	0.224	0.229	0.236	0.229	0.229	2.5				
Isopropylbenzene		3.061	3.316	3.102	3.325	3.639	3.599	3.341	7.2				
1,1,2,2-Tetrachloroethane		0.601	0.562	0.555	0.548	0.555	0.558	0.563	3.4				
1,2,3-Trichloropropane		0.498	0.445	0.418	0.405	0.408	0.397	0.428	8.8				
Bromobenzene		0.855	0.831	0.780	0.828	0.890	0.881	0.844	4.8				
n-propylbenzene		3.708	3.996	3.778	3.977	4.258	4.152	3.978	5.3				
2-Chlorotoluene		2.253	2.343	2.179	2.280	2.435	2.391	2.314	4.1				
1,3,5-Trimethylbenzene		2.467	2.757	2.614	2.800	2.978	2.898	2.752	6.8				
4-Chlorotoluene		2.313	2.501	2.270	2.380	2.521	2.445	2.405	4.2				
tert-Butylbenzene		2.246	2.382	2.263	2.494	2.725	2.701	2.469	8.5				
1,2,4-Trimethylbenzene		2.415	2.717	2.631	2.784	3.027	2.945	2.753	8				
sec-Butylbenzene		3.254	3.724	3.409	3.602	3.898	3.786	3.612	6.7				
p-Isopropyltoluene		2.717	3.016	2.869	3.094	3.386	3.321	3.067	8.4				
1,3-Dichlorobenzene		1.730	1.784	1.595	1.647	1.789	1.742	1.714	4.5				
1,4-Dichlorobenzene		1.821	1.726	1.605	1.620	1.733	1.682	1.698	4.7				

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG No.: Q1883
 Instrument ID: MSVOA_Y Calibration Date(s): 04/22/2025 04/22/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 13:39 16:15
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D		
		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
n-Butylbenzene	2.477	2.790	2.545	2.742	2.976	2.919	2.742	7.2
1,2-Dichlorobenzene	1.523	1.522	1.434	1.452	1.537	1.505	1.495	2.8
1,2-Dibromo-3-Chloropropane	0.097	0.091	0.094	0.084	0.085	0.087	0.090	5.9
1,2,4-Trichlorobenzene	0.747	0.800	0.785	0.823	0.970	0.959	0.847	11.1
Hexachlorobutadiene	0.522	0.545	0.483	0.480	0.555	0.544	0.521	6.3
1,2,3-Trichlorobenzene	0.671	0.673	0.684	0.707	0.829	0.825	0.731	10.2
1,2-Dichloroethane-d4	0.525	0.477	0.459	0.466	0.418	0.427	0.462	8.3
Dibromofluoromethane	0.335	0.341	0.330	0.330	0.319	0.327	0.330	2.3
Toluene-d8	1.220	1.260	1.211	1.276	1.244	1.261	1.245	2
4-Bromofluorobenzene	0.408	0.429	0.401	0.426	0.423	0.428	0.419	2.8
trans-1,4-Dichloro-2-butene	0.185	0.174	0.170	0.177	0.181	0.181	0.178	2.9

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.