

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1883

Client: Nobis Group

Analytical Method: SW8260D

Datafile : VY022147.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0505SBS01	Dichlorodifluoromethane	20	18.8	ug/Kg	94			29	149	
	Chloromethane	20	19.9	ug/Kg	100			50	136	
	Vinyl chloride	20	19.9	ug/Kg	100			56	135	
	Bromomethane	20	22.1	ug/Kg	111			53	143	
	Chloroethane	20	20.7	ug/Kg	104			59	139	
	Tetrahydrofuran	100	94.5	ug/Kg	95			61	135	
	Trichlorofluoromethane	20	20.5	ug/Kg	103			62	140	
	1,1,2-Trichlorotrifluoroethane	20	20.7	ug/Kg	104			66	136	
	1,1-Dichloroethene	20	20.0	ug/Kg	100			70	131	
	Acrylonitrile	100	94.5	ug/Kg	95			65	134	
	Acetone	100	93.4	ug/Kg	93			36	164	
	Carbon disulfide	20	19.2	ug/Kg	96			63	132	
	Methyl tert-butyl Ether	20	19.4	ug/Kg	97			73	125	
	Methylene Chloride	20	18.7	ug/Kg	94			70	128	
	trans-1,2-Dichloroethene	20	19.9	ug/Kg	100			74	125	
	1,1-Dichloroethane	20	20.1	ug/Kg	101			76	125	
	2-Butanone	100	92.1	ug/Kg	92			51	148	
	Carbon Tetrachloride	20	21.1	ug/Kg	106			70	135	
	2,2-Dichloropropane	20	21.2	ug/Kg	106			67	133	
	cis-1,2-Dichloroethene	20	19.8	ug/Kg	99			77	123	
	Chloroform	20	20.3	ug/Kg	102			78	123	
	1,1,1-Trichloroethane	20	20.8	ug/Kg	104			73	130	
	1,1-Dichloropropene	20	20.4	ug/Kg	102			76	125	
	Benzene	20	20.2	ug/Kg	101			77	121	
	1,2-Dichloroethane	20	21.0	ug/Kg	105			73	128	
	Trichloroethene	20	20.7	ug/Kg	104			77	123	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			76	123	
	Dibromomethane	20	20.1	ug/Kg	101			78	125	
	Bromodichloromethane	20	20.5	ug/Kg	103			75	127	
	4-Methyl-2-Pentanone	100	95.6	ug/Kg	96			65	135	
	Toluene	20	20.2	ug/Kg	101			77	121	
	t-1,3-Dichloropropene	20	20.2	ug/Kg	101			71	130	
	cis-1,3-Dichloropropene	20	20.0	ug/Kg	100			74	126	
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102			78	121	
	1,3-Dichloropropane	20	20.1	ug/Kg	101			77	121	
	2-Hexanone	100	92.9	ug/Kg	93			53	145	
	Dibromochloromethane	20	19.5	ug/Kg	98			74	126	
	1,2-Dibromoethane	20	19.6	ug/Kg	98			78	122	
	Tetrachloroethene	20	21.3	ug/Kg	106			73	128	
	Chlorobenzene	20	20.0	ug/Kg	100			79	120	
	1,1,1,2-Tetrachloroethane	20	20.0	ug/Kg	100			78	125	
	Ethyl Benzene	20	19.6	ug/Kg	98			76	122	
	m/p-Xylenes	40	40.3	ug/Kg	101			77	124	
	o-Xylene	20	19.9	ug/Kg	100			77	123	
	Styrene	20	19.8	ug/Kg	99			76	124	
	Bromoform	20	19.3	ug/Kg	97			67	132	
	Isopropylbenzene	20	19.5	ug/Kg	98			68	134	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/Kg	95			70	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1883

Client: Nobis Group

Analytical Method: SW8260D

Datafile : VY022147.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0505SBS01	1,2,3-Trichloropropane	20	20.2	ug/Kg	101			73	125	
	Bromobenzene	20	19.3	ug/Kg	97			78	121	
	N-propylbenzene	20	20.1	ug/Kg	101			73	125	
	2-Chlorotoluene	20	19.9	ug/Kg	100			75	122	
	1,3,5-Trimethylbenzene	20	20.1	ug/Kg	101			73	124	
	4-Chlorotoluene	20	20.2	ug/Kg	101			72	124	
	tert-Butylbenzene	20	19.2	ug/Kg	96			73	125	
	1,2,4-Trimethylbenzene	20	20.0	ug/Kg	100			75	123	
	Sec-butylbenzene	20	20.0	ug/Kg	100			73	126	
	p-Isopropyltoluene	20	20.0	ug/Kg	100			73	127	
	1,3-Dichlorobenzene	20	19.6	ug/Kg	98			77	121	
	1,4-Dichlorobenzene	20	19.5	ug/Kg	98			75	120	
	n-Butylbenzene	20	19.5	ug/Kg	98			70	128	
	1,2-Dichlorobenzene	20	19.4	ug/Kg	97			78	121	
	1,2-Dibromo-3-Chloropropane	20	19.6	ug/Kg	98			61	132	
	1,2,4-Trichlorobenzene	20	18.4	ug/Kg	92			67	129	
	Hexachlorobutadiene	20	19.5	ug/Kg	98			61	135	
	1,2,3-Trichlorobenzene	20	18.4	ug/Kg	92			66	130	
	trans-1,4-Dichloro-2-butene	20	19.3	ug/Kg	97			62	136	