

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03  
 Lab Code: CHEM Case No.: Q1883 SAS No.: Q1883 SDG NO.: Q1883  
 Lab File ID: VY022070.D Date Analyzed: 04/30/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:20  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                      | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD          | 233640        | 13.346 |  |  |  |  |
| UPPER LIMIT          | 467280        | 13.846 |  |  |  |  |
| LOWER LIMIT          | 116820        | 12.846 |  |  |  |  |
| EPA SAMPLE NO.       |               |        |  |  |  |  |
| OU4-PCS-TC-29-042325 | 194579        | 13.35  |  |  |  |  |
| OU4-PCS-TC-30-042325 | 200894        | 13.35  |  |  |  |  |
| OU4-PCS-TC-31-042325 | 189417        | 13.35  |  |  |  |  |
| OU4-PCS-TC-32-042325 | 231898        | 13.35  |  |  |  |  |
| OU4-VSL-18-042325    | 256467        | 13.35  |  |  |  |  |
| OU4-VSL-19-042325    | 197305        | 13.35  |  |  |  |  |
| VY0430SBL01          | 186896        | 13.35  |  |  |  |  |
| VY0430SBS01          | 222589        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.